

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR020824\
 Data File : PR065451.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Feb 2024 20:13
 Operator : AJ\MA
 Sample : P1286-11
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 BH6X5

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 02/12/2024
 Supervised By :Sohil Jodhani 02/12/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 09 22:07:15 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 31 14:15:44 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.700	3.001	87519949	94237166	13.462	16.231
2) SA Decachlor...	9.739	8.371	51162155	70173331	21.070	24.990
Target Compounds						
16) L4 AR-1242-1	4.939	4.146	2674178	1935467	17.522m	12.430m#
17) L4 AR-1242-2	4.964	4.165	4264748	2586800	19.032m	12.195m#
18) L4 AR-1242-3	5.029	4.349	3838679	2155557	26.914m	18.626 #
19) L4 AR-1242-4	5.133	4.442	3139267	3146024	27.586m	28.088
20) L4 AR-1242-5	5.933	5.002	9755439	10360944	80.099	72.549m
31) L7 AR-1260-1	6.684	5.736	10243699	12980879	37.833m	42.370m
32) L7 AR-1260-2	6.971	5.945	16052390	14728572	52.126	40.124
33) L7 AR-1260-3	7.368	6.106	10449326	10544212	45.499	30.498m#
34) L7 AR-1260-4	7.612	6.621	12867495	10721368	50.710m	37.507 #
35) L7 AR-1260-5	7.959	6.890	21026242	24915732	44.583	37.284

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR020824\
Data File : PR065451.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Feb 2024 20:13
Operator : AJ\MA
Sample : P1286-11
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
BH6X5

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/12/2024
Supervised By :Sohil Jodhani 02/12/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 09 22:07:15 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 31 14:15:44 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

