

Method Path : P:\HPCHEM1\ECD_R\Method\
 Method File : PR021018.M
 Title : GC EXTRACTABLES
 Last Update : Thu Feb 08 01:43:39 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR025181.D 1000 =PR025179.D 750 =PR025180.D
 250 =PR025182.D 50 =PR025183.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.749	1.292	1.740	1.788	1.851	1.684	E7	13.26
2) SA	Decachlorobiphenyl	0.972	1.234	1.060	1.052	1.343	1.132	E7	13.41
3) L1	AR-1016-1	4.322	3.282	4.299	4.575	4.308	4.157	E5	12.09
4) L1	AR-1016-2	6.022	4.794	6.119	6.518	6.956	6.082	E5	13.30
5) L1	AR-1016-3	5.035	4.062	5.172	5.392	5.412	5.015	E5	11.07
6) L1	AR-1016-4	4.792	3.943	4.956	5.444	5.717	4.970	E5	13.75
7) L1	AR-1016-5	4.770	3.871	4.852	5.424	5.493	4.882	E5	13.37
8) L2	AR-1221-1	1.872					1.872	E5	0.00
9) L2	AR-1221-2	1.433					1.433	E5	0.00
10) L2	AR-1221-3	4.267					4.267	E5	0.00
11) L3	AR-1232-1	3.464					3.464	E5	0.00
12) L3	AR-1232-2	1.653					1.653	E5	0.00
13) L3	AR-1232-3	2.477					2.477	E5	0.00
14) L3	AR-1232-4	1.949					1.949	E5	0.00
15) L3	AR-1232-5	1.995					1.995	E5	0.00
16) L4	AR-1242-1	2.732					2.732	E5	0.00
17) L4	AR-1242-2	6.296					6.296	E5	0.00
18) L4	AR-1242-3	3.958					3.958	E5	0.00
19) L4	AR-1242-4	3.264					3.264	E5	0.00
20) L4	AR-1242-5	3.263					3.263	E5	0.00
21) L5	AR-1248-1	5.389					5.389	E5	0.00
22) L5	AR-1248-2	5.981					5.981	E5	0.00
23) L5	AR-1248-3	5.260					5.260	E5	0.00
24) L5	AR-1248-4	6.890					6.890	E5	0.00
25) L5	AR-1248-5	6.971					6.971	E5	0.00
26) L6	AR-1254-1	4.066					4.066	E5	0.00
27) L6	AR-1254-2	1.055					1.055	E6	0.00
28) L6	AR-1254-3	1.127					1.127	E6	0.00
29) L6	AR-1254-4	8.180					8.180	E5	0.00
30) L6	AR-1254-5	8.264					8.264	E5	0.00
31) L7	AR-1260-1	8.376	7.941	9.205	9.293	9.980	8.959	E5	8.99
32) L7	AR-1260-2	0.949	0.951	1.065	1.053	1.100	1.023	E6	6.79
33) L7	AR-1260-3	6.723	7.060	7.663	7.510	7.846	7.360	E5	6.25
34) L7	AR-1260-4	7.624	8.296	8.775	8.449	9.238	8.476	E5	7.05
35) L7	AR-1260-5	1.433	1.689	1.683	1.546	1.641	1.598	E6	6.80
36) L8	AR-1262-1	6.184					6.184	E5	0.00
37) L8	AR-1262-2	6.982					6.982	E5	0.00
38) L8	AR-1262-3	9.317					9.317	E5	0.00
39) L8	AR-1262-4	8.518					8.518	E5	0.00
40) L8	AR-1262-5	1.549					1.549	E6	0.00
41) L9	AR-1268-1	1.661					1.661	E6	0.00
42) L9	AR-1268-2	1.531					1.531	E6	0.00
43) L9	AR-1268-3	1.247					1.247	E6	0.00
44) L9	AR-1268-4	5.751					5.751	E5	0.00
45) L9	AR-1268-5	3.889					3.889	E6	0.00

Signal #2 Calibration Files

500 =PR025181.D 1000 =PR025179.D 750 =PR025180.D
 250 =PR025182.D 50 =PR025183.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.966	2.192	2.930	3.041	3.223	2.870	E7	13.78
2) SA	Decachlorobiphenyl	1.862	2.359	1.981	1.929	2.445	2.115	E7	12.63
3) L1	AR-1016-1	0.904	0.672	0.870	0.998	1.070	0.903	E6	16.75
4) L1	AR-1016-2	7.249	5.746	7.375	7.796	8.622	7.358	E5	14.25
5) L1	AR-1016-3	7.997	6.241	7.936	9.050	9.352	8.115	E5	15.04

Method Path : P:\HPCHEM1\ECD_R\Method\
 Method File : PR021018.M
 Title : GC EXTRACTABLES
 Last Update : Thu Feb 08 01:43:39 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR025181.D 1000 =PR025179.D 750 =PR025180.D
 250 =PR025182.D 50 =PR025183.D

Compound			500	1000	750	250	50	Avg		%RSD
6)	L1	AR-1016-4	0.911	0.727	0.915	0.994	1.071	0.924	E6	13.85
7)	L1	AR-1016-5	0.872	0.727	0.896	0.955	1.022	0.894	E6	12.31
8)	L2	AR-1221-1	3.185					3.185	E5	0.00
9)	L2	AR-1221-2	2.341					2.341	E5	0.00
10)	L2	AR-1221-3	6.983					6.983	E5	0.00
11)	L3	AR-1232-1	5.767					5.767	E5	0.00
12)	L3	AR-1232-2	2.848					2.848	E5	0.00
13)	L3	AR-1232-3	2.146					2.146	E5	0.00
14)	L3	AR-1232-4	3.896					3.896	E5	0.00
15)	L3	AR-1232-5	3.486					3.486	E5	0.00
16)	L4	AR-1242-1	5.672					5.672	E5	0.00
17)	L4	AR-1242-2	7.793					7.793	E5	0.00
18)	L4	AR-1242-3	4.670					4.670	E5	0.00
19)	L4	AR-1242-4	3.640					3.640	E5	0.00
20)	L4	AR-1242-5	5.968					5.968	E5	0.00
21)	L5	AR-1248-1	7.772					7.772	E5	0.00
22)	L5	AR-1248-2	8.758					8.758	E5	0.00
23)	L5	AR-1248-3	1.108					1.108	E6	0.00
24)	L5	AR-1248-4	1.612					1.612	E6	0.00
25)	L5	AR-1248-5	1.134					1.134	E6	0.00
26)	L6	AR-1254-1	6.784					6.784	E5	0.00
27)	L6	AR-1254-2	1.480					1.480	E6	0.00
28)	L6	AR-1254-3	2.448					2.448	E6	0.00
29)	L6	AR-1254-4	1.514					1.514	E6	0.00
30)	L6	AR-1254-5	1.994					1.994	E6	0.00
31)	L7	AR-1260-1	1.613	1.520	1.764	1.785	1.999	1.736	E6	10.55
32)	L7	AR-1260-2	1.919	1.904	2.132	2.156	2.675	2.157	E6	14.47
33)	L7	AR-1260-3	2.078	2.186	2.376	2.300	2.437	2.275	E6	6.36
34)	L7	AR-1260-4	1.334	1.456	1.535	1.489	1.603	1.483	E6	6.74
35)	L7	AR-1260-5	3.151	3.750	3.697	3.399	3.666	3.533	E6	7.16
36)	L8	AR-1262-1	1.211					1.211	E6	0.00
37)	L8	AR-1262-2	1.420					1.420	E6	0.00
38)	L8	AR-1262-3	1.924					1.924	E6	0.00
39)	L8	AR-1262-4	1.693					1.693	E6	0.00
40)	L8	AR-1262-5	3.399					3.399	E6	0.00
41)	L9	AR-1268-1	3.525					3.525	E6	0.00
42)	L9	AR-1268-2	3.275					3.275	E6	0.00
43)	L9	AR-1268-3	2.645					2.645	E6	0.00
44)	L9	AR-1268-4	1.117					1.117	E6	0.00
45)	L9	AR-1268-5	8.184					8.184	E6	0.00

(#) = Out of Range