

Method Path : P:\HPCHEM1\ECD_R\Method\
 Method File : PR021018.M
 Title : GC EXTRACTABLES
 Last Update : Thu Feb 08 01:14:00 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR025181.D 1000 =PR025179.D 750 =PR025180.D
 250 =PR025182.D 50 =PR025183.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.749	1.292	1.740	1.788	1.851	1.684	E7	13.26
2) SA	Decachlorobiphenyl	0.972	1.234	1.060	1.052	1.343	1.132	E7	13.41
3) L1	AR-1016-1	4.322	3.282	4.299	4.575	4.308	4.157	E5	12.09
4) L1	AR-1016-2	6.022	4.794	6.119	6.518	6.956	6.082	E5	13.30
5) L1	AR-1016-3	5.035	4.062	5.172	5.392	5.412	5.015	E5	11.07
6) L1	AR-1016-4	4.792	3.943	4.956	5.444	5.717	4.970	E5	13.75
7) L1	AR-1016-5	4.770	3.871	4.852	5.424	5.493	4.882	E5	13.37
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1	8.376	7.941	9.205	9.293	9.980	8.959	E5	8.99
32) L7	AR-1260-2	0.949	0.951	1.065	1.053	1.100	1.023	E6	6.79
33) L7	AR-1260-3	6.723	7.060	7.663	7.510	7.846	7.360	E5	6.25
34) L7	AR-1260-4	7.624	8.296	8.775	8.449	9.238	8.476	E5	7.05
35) L7	AR-1260-5	1.433	1.689	1.683	1.546	1.641	1.598	E6	6.80
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =PR025181.D 1000 =PR025179.D 750 =PR025180.D
 250 =PR025182.D 50 =PR025183.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.966	2.192	2.930	3.041	3.223	2.870	E7	13.78
2) SA	Decachlorobiphenyl	1.862	2.359	1.981	1.929	2.445	2.115	E7	12.63
3) L1	AR-1016-1	0.904	0.672	0.870	0.998	1.070	0.903	E6	16.75
4) L1	AR-1016-2	7.249	5.746	7.375	7.796	8.622	7.358	E5	14.25
5) L1	AR-1016-3	7.997	6.241	7.936	9.050	9.352	8.115	E5	15.04

Method Path : P:\HPCHEM1\ECD_R\Method\
 Method File : PR021018.M
 Title : GC EXTRACTABLES
 Last Update : Thu Feb 08 01:14:00 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR025181.D 1000 =PR025179.D 750 =PR025180.D
 250 =PR025182.D 50 =PR025183.D

Compound			500	1000	750	250	50	Avg	%RSD

6)	L1	AR-1016-4	0.911	0.727	0.915	0.994	1.071	0.924 E6	13.85
7)	L1	AR-1016-5	0.872	0.727	0.896	0.955	1.022	0.894 E6	12.31
8)	L2	AR-1221-1						0.000	-1.00
9)	L2	AR-1221-2						0.000	-1.00
10)	L2	AR-1221-3						0.000	-1.00
11)	L3	AR-1232-1						0.000	-1.00
12)	L3	AR-1232-2						0.000	-1.00
13)	L3	AR-1232-3						0.000	-1.00
14)	L3	AR-1232-4						0.000	-1.00
15)	L3	AR-1232-5						0.000	-1.00
16)	L4	AR-1242-1						0.000	-1.00
17)	L4	AR-1242-2						0.000	-1.00
18)	L4	AR-1242-3						0.000	-1.00
19)	L4	AR-1242-4						0.000	-1.00
20)	L4	AR-1242-5						0.000	-1.00
21)	L5	AR-1248-1						0.000	-1.00
22)	L5	AR-1248-2						0.000	-1.00
23)	L5	AR-1248-3						0.000	-1.00
24)	L5	AR-1248-4						0.000	-1.00
25)	L5	AR-1248-5						0.000	-1.00
26)	L6	AR-1254-1						0.000	-1.00
27)	L6	AR-1254-2						0.000	-1.00
28)	L6	AR-1254-3						0.000	-1.00
29)	L6	AR-1254-4						0.000	-1.00
30)	L6	AR-1254-5						0.000	-1.00
31)	L7	AR-1260-1	1.613	1.520	1.764	1.785	1.999	1.736 E6	10.55
32)	L7	AR-1260-2	1.919	1.904	2.132	2.156	2.675	2.157 E6	14.47
33)	L7	AR-1260-3	2.078	2.186	2.376	2.300	2.437	2.275 E6	6.36
34)	L7	AR-1260-4	1.334	1.456	1.535	1.489	1.603	1.483 E6	6.74
35)	L7	AR-1260-5	3.151	3.750	3.697	3.399	3.666	3.533 E6	7.16
36)	L8	AR-1262-1						0.000	-1.00
37)	L8	AR-1262-2						0.000	-1.00
38)	L8	AR-1262-3						0.000	-1.00
39)	L8	AR-1262-4						0.000	-1.00
40)	L8	AR-1262-5						0.000	-1.00
41)	L9	AR-1268-1						0.000	-1.00
42)	L9	AR-1268-2						0.000	-1.00
43)	L9	AR-1268-3						0.000	-1.00
44)	L9	AR-1268-4						0.000	-1.00
45)	L9	AR-1268-5						0.000	-1.00

 (#) = Out of Range