

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR021023\
 Data File : PR059483.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Feb 2023 13:28
 Operator : YP\AJ
 Sample : PB150764BS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 ALCS764

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 10 21:41:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 27 08:44:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.693	2.908	58543074	87673578	21.581	22.049
2) SA Decachlor...	9.667	8.144	83763511	132.5E6	40.049	41.525
Target Compounds						
3) L1 AR-1016-1	4.928	4.019	9617516	13219382	111.033	95.211
4) L1 AR-1016-2	4.950	4.037	13180299	18882190	110.888	98.004
5) L1 AR-1016-3	5.015	4.215	8925265	10019873	116.694	98.698
6) L1 AR-1016-4	5.119	4.265	7008298	8556978	112.642	111.415
7) L1 AR-1016-5	5.437	4.481	6708002	10685054	109.129	103.978
31) L7 AR-1260-1	6.660	5.568	12076999	20791186	103.139	97.303
32) L7 AR-1260-2	6.945	5.772	13735993	24344808	98.987	94.671
33) L7 AR-1260-3	7.337	5.930	8914758	22730285	86.328	96.464
34) L7 AR-1260-4	7.582	6.435	11084182	16502199	94.625	85.405
35) L7 AR-1260-5	7.921	6.699	19265065	37225717	84.776	83.032

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR021023\
Data File : PR059483.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Feb 2023 13:28
Operator : YP\AJ
Sample : PB150764BS
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
ALCS764

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 10 21:41:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 27 08:44:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

