

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR021025\
 Data File : PR070282.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Feb 2025 23:29
 Operator : AJ\MA
 Sample : Q1168-07
 Misc : AR1232 LOD 40 PPB
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 LOD-MDL-WATER-01-QT1-2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 11 00:53:04 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 10 14:13:17 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.624	2.956	18969776	98080700	16.800	16.820
2)	SA Decachlor...	9.509	8.223	13757686	77498164	19.245	18.928
Target Compounds							
3)	L1 AR-1016-1	4.844	4.073	422460	2460135	14.234	15.596
4)	L1 AR-1016-2	4.867	4.091	823977	3743285	17.285	15.713
5)	L1 AR-1016-3	4.931	4.271	558542	1748806	18.178	13.934
6)	L1 AR-1016-4	5.033	4.323	385844	1361039	15.870	12.384
7)	L1 AR-1016-5	5.345	4.541	351924	1759242	14.689	12.897
8)	L2 AR-1221-1	3.839	3.179	549196	1354548	42.072	19.789 #
9)	L2 AR-1221-2	3.936	3.262	465440	2432164	47.673	47.500
10)	L2 AR-1221-3	4.012	3.342	881599	4132789	28.827	26.545
11)	L3 AR-1232-1	4.012	3.342	881599	4132789	37.172	34.562
12)	L3 AR-1232-2	4.558	4.091	477129	3743285	38.175	33.720
13)	L3 AR-1232-3	4.867	4.271	823977	1748806	38.105	30.299
14)	L3 AR-1232-4	5.033	4.364	385844	1682013	35.407	31.696
15)	L3 AR-1232-5	5.183	4.541	219091	1759242	22.820	30.747 #
16)	L4 AR-1242-1	4.844	4.073	422460	2460135	17.593	19.023
17)	L4 AR-1242-2	4.867	4.091	823977	3743285	21.243	19.246
18)	L4 AR-1242-3	4.931	4.271	558542	1748806	22.268	16.805
19)	L4 AR-1242-4	5.033	4.364	385844	1682013	19.860	16.411
20)	L4 AR-1242-5	5.817	4.912	319724	2192460	16.886	19.133
21)	L5 AR-1248-1	4.844	4.073	422460	2460135	23.254	25.500
22)	L5 AR-1248-2	5.136	4.323	272001	1361039	9.806	9.381
23)	L5 AR-1248-3	5.345	4.364	351924	1682013	11.176	11.360
24)	L5 AR-1248-4	5.781	4.541	253222	1759242	8.415	10.009
25)	L5 AR-1248-5	5.817	4.954	319724	1907097	10.061	12.701 #
26)	L6 AR-1254-1	5.750	4.912	241363	2192460	6.624	8.675 #
27)	L6 AR-1254-2	5.985	5.071	227190	1085053	4.198	4.980
28)	L6 AR-1254-3	6.381	5.495	191793	688714	3.589	2.158 #
29)	L6 AR-1254-4	6.680	5.744	621608	465118	19.454	2.890 #
30)	L6 AR-1254-5	0.000	6.189	0	341989	N.D.	1.315 #
31)	L7 AR-1260-1	0.000	5.640	0	574668	N.D.	2.612 #
32)	L7 AR-1260-2	0.000	5.843	0	180130	N.D.	0.732 #
33)	L7 AR-1260-3	0.000	6.005	0	569493	N.D.	2.390 #
34)	L7 AR-1260-4	7.472	6.505	90659	208440	2.276	1.058 #
35)	L7 AR-1260-5	0.000	6.774	0	401416	N.D.	0.976 #
36)	L8 AR-1262-1	7.472	6.227	90659	297887	1.797	0.968 #
37)	L8 AR-1262-2	0.000	6.505	0	208440	N.D.	0.772 #
38)	L8 AR-1262-3	0.000	7.083	0	356429	N.D.	1.817 #
39)	L8 AR-1262-4	0.000	7.153	0	338263	N.D.	0.939 #
40)	L8 AR-1262-5	0.000	7.689	0	171536	N.D.	1.125 #
41)	L9 AR-1268-1	0.000	7.083	0	356429	N.D.	0.624 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR021025\
 Data File : PR070282.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Feb 2025 23:29
 Operator : AJ\MA
 Sample : Q1168-07
 Misc : AR1232 LOD 40 PPB
 ALS Vial : 44 Sample Multiplier: 1


Instrument :
 ECD_R
 ClientSampleId :
 LOD-MDL-WATER-01-QT1-2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 11 00:53:04 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 10 14:13:17 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	0.000	7.153	0	338263	N.D.	0.646 #
43) L9	AR-1268-3	0.000	7.367	0	397684	N.D.	0.813 #
44) L9	AR-1268-4	0.000	7.689	0	171536	N.D.	1.002 #
45) L9	AR-1268-5	0.000	7.981	0	478750	N.D.	0.345 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.



#1 Tetrachloro-m-xylene


R.T.: 3.624 min
Delta R.T.: 0.000 min
Response: 18969776
Conc: 16.80 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#1 Tetrachloro-m-xylene

R.T.: 2.956 min
Delta R.T.: 0.000 min
Response: 98080700
Conc: 16.82 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.509 min
Delta R.T.: -0.003 min
Response: 13757686
Conc: 19.24 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.223 min
Delta R.T.: -0.003 min
Response: 77498164
Conc: 18.93 ng/ml



#3 AR-1016-1


R.T.: 4.844 min
Delta R.T.: 0.000 min
Response: 422460
Conc: 14.23 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#3 AR-1016-1

R.T.: 4.073 min
Delta R.T.: -0.001 min
Response: 2460135
Conc: 15.60 ng/ml




#4 AR-1016-2

R.T.: 4.867 min
Delta R.T.: 0.001 min
Response: 823977
Conc: 17.29 ng/ml



#4 AR-1016-2


R.T.: 4.091 min
Delta R.T.: 0.000 min
Response: 3743285
Conc: 15.71 ng/ml



#5 AR-1016-3


R.T.: 4.931 min
Delta R.T.: 0.001 min
Response: 558542
Conc: 18.18 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#5 AR-1016-3

R.T.: 4.271 min
Delta R.T.: -0.001 min
Response: 1748806
Conc: 13.93 ng/ml




#6 AR-1016-4

R.T.: 5.033 min
Delta R.T.: 0.000 min
Response: 385844
Conc: 15.87 ng/ml



#6 AR-1016-4


R.T.: 4.323 min
Delta R.T.: 0.000 min
Response: 1361039
Conc: 12.38 ng/ml



#7 AR-1016-5


R.T.: 5.345 min
Delta R.T.: -0.002 min
Response: 351924
Conc: 14.69 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#7 AR-1016-5

R.T.: 4.541 min
Delta R.T.: 0.000 min
Response: 1759242
Conc: 12.90 ng/ml




#8 AR-1221-1

R.T.: 3.839 min
Delta R.T.: -0.007 min
Response: 549196
Conc: 42.07 ng/ml



#8 AR-1221-1


R.T.: 3.179 min
Delta R.T.: 0.000 min
Response: 1354548
Conc: 19.79 ng/ml



#9 AR-1221-2


R.T.: 3.936 min
Delta R.T.: 0.001 min
Response: 465440
Conc: 47.67 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#9 AR-1221-2

R.T.: 3.262 min
Delta R.T.: -0.003 min
Response: 2432164
Conc: 47.50 ng/ml




#10 AR-1221-3

R.T.: 4.012 min
Delta R.T.: 0.000 min
Response: 881599
Conc: 28.83 ng/ml



#10 AR-1221-3


R.T.: 3.342 min
Delta R.T.: 0.000 min
Response: 4132789
Conc: 26.55 ng/ml



#11 AR-1232-1


R.T.: 4.012 min
Delta R.T.: 0.000 min
Response: 881599
Conc: 37.17 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#11 AR-1232-1

R.T.: 3.342 min
Delta R.T.: 0.000 min
Response: 4132789
Conc: 34.56 ng/ml




#12 AR-1232-2

R.T.: 4.558 min
Delta R.T.: -0.001 min
Response: 477129
Conc: 38.18 ng/ml



#12 AR-1232-2


R.T.: 4.091 min
Delta R.T.: -0.001 min
Response: 3743285
Conc: 33.72 ng/ml



#13 AR-1232-3


R.T.: 4.867 min
Delta R.T.: 0.001 min
Response: 823977
Conc: 38.10 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#13 AR-1232-3

R.T.: 4.271 min
Delta R.T.: -0.001 min
Response: 1748806
Conc: 30.30 ng/ml




#14 AR-1232-4

R.T.: 5.033 min
Delta R.T.: 0.000 min
Response: 385844
Conc: 35.41 ng/ml



#14 AR-1232-4


R.T.: 4.364 min
Delta R.T.: -0.001 min
Response: 1682013
Conc: 31.70 ng/ml



#15 AR-1232-5


R.T.: 5.183 min
Delta R.T.: 0.000 min
Response: 219091
Conc: 22.82 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#15 AR-1232-5

R.T.: 4.541 min
Delta R.T.: 0.000 min
Response: 1759242
Conc: 30.75 ng/ml




#16 AR-1242-1

R.T.: 4.844 min
Delta R.T.: 0.000 min
Response: 422460
Conc: 17.59 ng/ml



#16 AR-1242-1


R.T.: 4.073 min
Delta R.T.: 0.000 min
Response: 2460135
Conc: 19.02 ng/ml



#17 AR-1242-2


R.T.: 4.867 min
Delta R.T.: 0.002 min
Response: 823977
Conc: 21.24 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#17 AR-1242-2

R.T.: 4.091 min
Delta R.T.: 0.000 min
Response: 3743285
Conc: 19.25 ng/ml




#18 AR-1242-3

R.T.: 4.931 min
Delta R.T.: 0.002 min
Response: 558542
Conc: 22.27 ng/ml



#18 AR-1242-3


R.T.: 4.271 min
Delta R.T.: 0.000 min
Response: 1748806
Conc: 16.81 ng/ml



#19 AR-1242-4


R.T.: 5.033 min
Delta R.T.: 0.000 min
Response: 385844
Conc: 19.86 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#19 AR-1242-4

R.T.: 4.364 min
Delta R.T.: 0.000 min
Response: 1682013
Conc: 16.41 ng/ml




#20 AR-1242-5

R.T.: 5.817 min
Delta R.T.: -0.002 min
Response: 319724
Conc: 16.89 ng/ml



#20 AR-1242-5


R.T.: 4.912 min
Delta R.T.: 0.000 min
Response: 2192460
Conc: 19.13 ng/ml



#21 AR-1248-1


R.T.: 4.844 min
Delta R.T.: 0.000 min
Response: 422460
Conc: 23.25 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#21 AR-1248-1

R.T.: 4.073 min
Delta R.T.: 0.000 min
Response: 2460135
Conc: 25.50 ng/ml




#22 AR-1248-2

R.T.: 5.136 min
Delta R.T.: 0.001 min
Response: 272001
Conc: 9.81 ng/ml



#22 AR-1248-2


R.T.: 4.323 min
Delta R.T.: 0.000 min
Response: 1361039
Conc: 9.38 ng/ml



#23 AR-1248-3


R.T.: 5.345 min
Delta R.T.: -0.002 min
Response: 351924
Conc: 11.18 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#23 AR-1248-3

R.T.: 4.364 min
Delta R.T.: 0.000 min
Response: 1682013
Conc: 11.36 ng/ml




#24 AR-1248-4

R.T.: 5.781 min
Delta R.T.: 0.003 min
Response: 253222
Conc: 8.42 ng/ml



#24 AR-1248-4


R.T.: 4.541 min
Delta R.T.: 0.000 min
Response: 1759242
Conc: 10.01 ng/ml



#25 AR-1248-5


R.T.: 5.817 min
Delta R.T.: -0.002 min
Response: 319724
Conc: 10.06 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#25 AR-1248-5

R.T.: 4.954 min
Delta R.T.: 0.000 min
Response: 1907097
Conc: 12.70 ng/ml




#26 AR-1254-1

R.T.: 5.750 min
Delta R.T.: 0.000 min
Response: 241363
Conc: 6.62 ng/ml



#26 AR-1254-1


R.T.: 4.912 min
Delta R.T.: 0.001 min
Response: 2192460
Conc: 8.68 ng/ml



#27 AR-1254-2


R.T.: 5.985 min
Delta R.T.: -0.001 min
Response: 227190
Conc: 4.20 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#27 AR-1254-2

R.T.: 5.071 min
Delta R.T.: -0.001 min
Response: 1085053
Conc: 4.98 ng/ml




#28 AR-1254-3

R.T.: 6.381 min
Delta R.T.: 0.002 min
Response: 191793
Conc: 3.59 ng/ml



#28 AR-1254-3


R.T.: 5.495 min
Delta R.T.: 0.000 min
Response: 688714
Conc: 2.16 ng/ml



#29 AR-1254-4


R.T.: 6.680 min
Delta R.T.: -0.011 min
Response: 621608
Conc: 19.45 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#29 AR-1254-4

R.T.: 5.744 min
Delta R.T.: 0.000 min
Response: 465118
Conc: 2.89 ng/ml




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 7.147 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 6.189 min
Delta R.T.: 0.000 min
Response: 341989
Conc: 1.32 ng/ml



#31 AR-1260-1


R.T.: 0.000 min
Exp R.T.: 6.559 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#31 AR-1260-1

R.T.: 5.640 min
Delta R.T.: 0.003 min
Response: 574668
Conc: 2.61 ng/ml




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T.: 6.842 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 5.843 min
Delta R.T.: 0.000 min
Response: 180130
Conc: 0.73 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T.: 7.233 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#33 AR-1260-3

R.T.: 6.005 min
Delta R.T.: 0.004 min
Response: 569493
Conc: 2.39 ng/ml




#34 AR-1260-4

R.T.: 7.472 min
Delta R.T.: -0.004 min
Response: 90659
Conc: 2.28 ng/ml



#34 AR-1260-4


R.T.: 6.505 min
Delta R.T.: -0.002 min
Response: 208440
Conc: 1.06 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T.: 7.816 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#35 AR-1260-5

R.T.: 6.774 min
Delta R.T.: 0.000 min
Response: 401416
Conc: 0.98 ng/ml




#36 AR-1262-1

R.T.: 7.472 min
Delta R.T.: -0.004 min
Response: 90659
Conc: 1.80 ng/ml



#36 AR-1262-1


R.T.: 6.227 min
Delta R.T.: -0.006 min
Response: 297887
Conc: 0.97 ng/ml



#37 AR-1262-2


R.T.: 0.000 min
Exp R.T.: 7.815 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#37 AR-1262-2

R.T.: 6.505 min
Delta R.T.: -0.002 min
Response: 208440
Conc: 0.77 ng/ml




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T.: 8.122 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 7.083 min
Delta R.T.: 0.003 min
Response: 356429
Conc: 1.82 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T.: 8.206 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#39 AR-1262-4

R.T.: 7.153 min
Delta R.T.: 0.006 min
Response: 338263
Conc: 0.94 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T.: 8.828 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 7.689 min
Delta R.T.: 0.002 min
Response: 171536
Conc: 1.12 ng/ml



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T.: 8.119 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#41 AR-1268-1

R.T.: 7.083 min
Delta R.T.: 0.005 min
Response: 356429
Conc: 0.62 ng/ml




#42 AR-1268-2

R.T.: 0.000 min
Exp R.T.: 8.210 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 7.153 min
Delta R.T.: 0.005 min
Response: 338263
Conc: 0.65 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 8.424 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025




#43 AR-1268-3

R.T.: 7.367 min
Delta R.T.: -0.002 min
Response: 397684
Conc: 0.81 ng/ml




#44 AR-1268-4

R.T.: 0.000 min
Exp R.T.: 8.828 min
Response: 0
Conc: N.D.



#44 AR-1268-4


R.T.: 7.689 min
Delta R.T.: 0.003 min
Response: 171536
Conc: 1.00 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 9.207 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2025



#45 AR-1268-5

R.T.: 7.981 min
Delta R.T.: 0.000 min
Response: 478750
Conc: 0.35 ng/ml