

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR021519\
 Data File : PR035471.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Feb 2019 08:11
 Operator : SM\SJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 15 22:31:53 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR020119CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Feb 02 01:45:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds									
20)	L4	AR-1242-5	6.459f	0.000	1294004	0	25.578	N.D.	#
24)	L5	AR-1248-4	6.459	0.000	1294004	0	16.454	N.D.	#
25)	L5	AR-1248-5	6.459f	0.000	1294004	0	17.405	N.D.	#
26)	L6	AR-1254-1	6.459	0.000	1294004	0	11.655	N.D.	#
27)	L6	AR-1254-2	6.654	0.000	1965839	0	11.067	N.D.	#
28)	L6	AR-1254-3	7.063f	0.000	1363599	0	7.147	N.D.	#
29)	L6	AR-1254-4	7.344f	0.000	1160184	0	7.204	N.D.	#
30)	L6	AR-1254-5	7.776f	0.000	2966590	0	17.981	N.D.	#
31)	L7	AR-1260-1	7.186	0.000	818055	0	4.756	N.D.	#
33)	L7	AR-1260-3	7.776f	0.000	2966590	0	11.139	N.D.	#
36)	L8	AR-1262-1	7.815	0.000	1167602	0	6.306	N.D.	#
45)	L9	AR-1268-5	9.771	0.000	134586	0	0.113	N.D.	#


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR021519\
Data File : PR035471.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Feb 2019 08:11
Operator : SM\SJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 15 22:31:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR020119CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Feb 02 01:45:00 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#20 AR-1242-5


R.T.: 6.459 min
Delta R.T.: -0.046 min
Response: 1294004
Conc: 25.58 ng/ml

Instrument :
ECD_R
ClientSampleId :
HEXANE




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T. : 5.344 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 6.459 min
Delta R.T.: -0.008 min
Response: 1294004
Conc: 16.45 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 5.000 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.459 min
Delta R.T.: -0.046 min
Response: 1294004
Conc: 17.40 ng/ml

Instrument :
ECD_R
ClientSampleId :
HEXANE




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T. : 5.383 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 6.459 min
Delta R.T.: 0.019 min
Response: 1294004
Conc: 11.66 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T. : 5.344 min
Response: 0
Conc: N.D.



#27 AR-1254-2


R.T.: 6.654 min
Delta R.T.: -0.002 min
Response: 1965839
Conc: 11.07 ng/ml

Instrument :
ECD_R
ClientSampleId :
HEXANE




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T. : 5.488 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 7.063 min
Delta R.T.: 0.041 min
Response: 1363599
Conc: 7.15 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 5.884 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 7.344 min
Delta R.T.: 0.038 min
Response: 1160184
Conc: 7.20 ng/ml

Instrument :
ECD_R
ClientSampleId :
HEXANE




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.110 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 7.776 min
Delta R.T.: 0.053 min
Response: 2966590
Conc: 17.98 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T. : 6.521 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 7.186 min
Delta R.T.: 0.001 min
Response: 818055
Conc: 4.76 ng/ml

Instrument :
ECD_R
ClientSampleId :
HEXANE




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T. : 6.012 min
Response: 0
Conc: N.D.




#33 AR-1260-3

R.T.: 7.776 min
Delta R.T.: 0.053 min
Response: 2966590
Conc: 11.14 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T. : 6.348 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 7.815 min
Delta R.T.: 0.019 min
Response: 1167602
Conc: 6.31 ng/ml

Instrument :
ECD_R
ClientSampleId :
HEXANE




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T. : 6.556 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.771 min
Delta R.T.: -0.006 min
Response: 134586
Conc: 0.11 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 8.155 min
Response: 0
Conc: N.D.