

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR021522CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Feb 16 02:41:16 2022
 Response Via : Initial Calibration

Calibration Files

400 =PR053530.D 1600 =PR053532.D 800 =PR053531.D
 200 =PR053529.D 100 =PR053528.D

	Compound	400	1600	800	200	100	Avg		%RSD
1)	SA Tetrachloro-m-xylene	2.061	1.816	1.896	2.100	2.157	2.006	E6	7.17
2)	SA Decachlorobiphenyl	1.645	1.473	1.553	1.780	1.762	1.642	E6	8.05
3)	L1 AR-1016-1						0.000		-1.00
4)	L1 AR-1016-2						0.000		-1.00
5)	L1 AR-1016-3						0.000		-1.00
6)	L1 AR-1016-4						0.000		-1.00
7)	L1 AR-1016-5						0.000		-1.00
8)	L2 AR-1221-1	2.474	2.015	2.189	2.591	2.685	2.391	E4	11.75
9)	L2 AR-1221-2	1.729	1.398	1.585	1.825	1.897	1.687	E4	11.81
10)	L2 AR-1221-3	5.533	4.516	5.063	6.051	6.240	5.481	E4	12.93
11)	L3 AR-1232-1						0.000		-1.00
12)	L3 AR-1232-2						0.000		-1.00
13)	L3 AR-1232-3						0.000		-1.00
14)	L3 AR-1232-4						0.000		-1.00
15)	L3 AR-1232-5						0.000		-1.00
16)	L4 AR-1242-1						0.000		-1.00
17)	L4 AR-1242-2						0.000		-1.00
18)	L4 AR-1242-3						0.000		-1.00
19)	L4 AR-1242-4						0.000		-1.00
20)	L4 AR-1242-5						0.000		-1.00
21)	L5 AR-1248-1						0.000		-1.00
22)	L5 AR-1248-2						0.000		-1.00
23)	L5 AR-1248-3						0.000		-1.00
24)	L5 AR-1248-4						0.000		-1.00
25)	L5 AR-1248-5						0.000		-1.00
26)	L6 AR-1254-1						0.000		-1.00
27)	L6 AR-1254-2						0.000		-1.00
28)	L6 AR-1254-3						0.000		-1.00
29)	L6 AR-1254-4						0.000		-1.00
30)	L6 AR-1254-5						0.000		-1.00
31)	L7 AR-1260-1						0.000		-1.00
32)	L7 AR-1260-2						0.000		-1.00
33)	L7 AR-1260-3						0.000		-1.00
34)	L7 AR-1260-4						0.000		-1.00
35)	L7 AR-1260-5						0.000		-1.00
36)	L8 AR-1262-1						0.000		-1.00
37)	L8 AR-1262-2						0.000		-1.00
38)	L8 AR-1262-3						0.000		-1.00
39)	L8 AR-1262-4						0.000		-1.00
40)	L8 AR-1262-5						0.000		-1.00
41)	L9 AR-1268-1						0.000		-1.00
42)	L9 AR-1268-2						0.000		-1.00
43)	L9 AR-1268-3						0.000		-1.00
44)	L9 AR-1268-4						0.000		-1.00
45)	L9 AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR053530.D 1600 =PR053532.D 800 =PR053531.D
 200 =PR053529.D 100 =PR053528.D

	Compound	400	1600	800	200	100	Avg		%RSD
1)	SA Tetrachloro-m-xylene	1.186	1.096	1.124	1.236	1.239	1.176	E6	5.51
2)	SA Decachlorobiphenyl	1.303	1.168	1.227	1.396	1.473	1.313	E6	9.39
3)	L1 AR-1016-1						0.000		-1.00
4)	L1 AR-1016-2						0.000		-1.00
5)	L1 AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR021522CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Feb 16 02:41:16 2022
 Response Via : Initial Calibration

Calibration Files

400 =PR053530.D 1600 =PR053532.D 800 =PR053531.D
 200 =PR053529.D 100 =PR053528.D

	Compound	400	1600	800	200	100	Avg	%RSD

6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1	1.568	1.328	1.458	1.618	1.742	1.543 E4	10.21
9) L2	AR-1221-2	1.101	0.939	1.024	1.145	1.195	1.081 E4	9.35
10) L2	AR-1221-3	3.755	3.089	3.376	4.064	4.194	3.696 E4	12.53
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

 (#) = Out of Range