

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR021522CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Feb 16 03:56:47 2022
 Response Via : Initial Calibration

Calibration Files

400 =PR053545.D 1600 =PR053547.D 800 =PR053546.D
 200 =PR053544.D 100 =PR053543.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.936	1.881	1.957	2.048	2.153	1.995	E6	5.35
2) SA	Decachlorobiphenyl	1.741	1.507	1.621	1.854	1.914	1.728	E6	9.65
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1	4.048	3.389	3.732	4.328	4.506	4.001	E4	11.26
22) L5	AR-1248-2	5.545	4.584	5.088	6.088	6.429	5.547	E4	13.40
23) L5	AR-1248-3	6.548	5.490	6.005	7.074	7.412	6.506	E4	11.98
24) L5	AR-1248-4	7.339	6.382	6.830	7.947	8.323	7.364	E4	10.76
25) L5	AR-1248-5	7.162	6.232	6.667	7.783	8.248	7.218	E4	11.29
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR053545.D 1600 =PR053547.D 800 =PR053546.D
 200 =PR053544.D 100 =PR053543.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.297	1.196	1.262	1.357	1.343	1.291	E6	5.05
2) SA	Decachlorobiphenyl	1.370	1.210	1.277	1.473	1.545	1.375	E6	9.99
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR021522CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Feb 16 03:56:47 2022
 Response Via : Initial Calibration

Calibration Files

400 =PR053545.D 1600 =PR053547.D 800 =PR053546.D
 200 =PR053544.D 100 =PR053543.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1	2.728	2.280	2.517	2.917	3.120	2.712 E4	12.13
22) L5	AR-1248-2	4.316	3.547	3.940	4.744	4.959	4.301 E4	13.40
23) L5	AR-1248-3	4.255	3.481	3.865	4.601	4.750	4.190 E4	12.49
24) L5	AR-1248-4	5.104	4.316	4.741	5.552	5.676	5.078 E4	11.13
25) L5	AR-1248-5	4.968	4.329	4.679	5.424	5.258	4.932 E4	8.94
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range