

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR021522CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Feb 16 04:28:13 2022
 Response Via : Initial Calibration

Calibration Files

400 =PR053550.D 1600 =PR053552.D 800 =PR053551.D
 200 =PR053549.D 100 =PR053548.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	1.986	1.867	1.908	2.030	2.112	1.980	E6	4.92
2)	SA	Decachlorobiphenyl	1.738	1.500	1.614	1.821	1.861	1.707	E6	8.75
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1	7.161	6.113	6.603	7.611	8.232	7.144	E4	11.62
27)	L6	AR-1254-2	1.121	0.958	1.035	1.205	1.312	1.126	E5	12.33
28)	L6	AR-1254-3	1.178	1.039	1.102	1.239	1.313	1.174	E5	9.24
29)	L6	AR-1254-4	0.910	0.789	0.844	0.959	1.035	0.908	E5	10.60
30)	L6	AR-1254-5	0.975	0.850	0.908	1.020	1.094	0.969	E5	9.79
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR053550.D 1600 =PR053552.D 800 =PR053551.D
 200 =PR053549.D 100 =PR053548.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.266	1.166	1.225	1.310	1.323	1.258	E6	5.09
2) SA	Decachlorobiphenyl	1.349	1.191	1.262	1.434	1.512	1.350	E6	9.54
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR021522CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Feb 16 04:28:13 2022
 Response Via : Initial Calibration

Calibration Files

400 =PR053550.D 1600 =PR053552.D 800 =PR053551.D
 200 =PR053549.D 100 =PR053548.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	7.524	6.486	6.945	8.043	8.511	7.502 E4	10.86
27) L6	AR-1254-2	6.592	5.598	6.046	7.106	7.525	6.574 E4	11.83
28) L6	AR-1254-3	1.036	0.913	0.963	1.090	1.153	1.031 E5	9.32
29) L6	AR-1254-4	6.604	5.788	6.183	7.027	7.313	6.583 E4	9.38
30) L6	AR-1254-5	0.957	0.837	0.895	0.984	1.033	0.941 E5	8.12
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range