

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR021522CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Feb 16 05:10:26 2022
 Response Via : Initial Calibration

Calibration Files

400 =PR053560.D 1600 =PR053562.D 800 =PR053561.D
 200 =PR053559.D 100 =PR053558.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.970	1.861	1.916	2.055	2.122	1.985	E6	5.27
2) SA	Decachlorobiphenyl	2.952	2.630	2.721	3.135	3.308	2.949	E6	9.54
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1	2.405	2.183	2.295	2.560	2.758	2.440	E5	9.26
42) L9	AR-1268-2	2.194	2.045	2.107	2.330	2.534	2.242	E5	8.70
43) L9	AR-1268-3	1.859	1.724	1.789	1.975	2.105	1.890	E5	8.04
44) L9	AR-1268-4	8.437	7.631	7.857	8.689	8.966	8.316	E4	6.74
45) L9	AR-1268-5	6.195	5.893	5.888	6.400	6.669	6.209	E5	5.41

Signal #2 Calibration Files

400 =PR053560.D 1600 =PR053562.D 800 =PR053561.D
 200 =PR053559.D 100 =PR053558.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.259	1.171	1.217	1.304	1.335	1.257	E6	5.24
2) SA	Decachlorobiphenyl	2.290	2.074	2.165	2.421	2.587	2.307	E6	8.84
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR021522CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Feb 16 05:10:26 2022
 Response Via : Initial Calibration

Calibration Files

400 =PR053560.D 1600 =PR053562.D 800 =PR053561.D
 200 =PR053559.D 100 =PR053558.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	2.079	1.946	1.982	2.197	2.305	2.102 E5	7.10
42) L9	AR-1268-2	1.884	1.766	1.806	1.982	2.088	1.905 E5	6.88
43) L9	AR-1268-3	1.613	1.505	1.538	1.701	1.788	1.629 E5	7.17
44) L9	AR-1268-4	7.083	6.339	6.647	7.575	8.098	7.149 E4	9.88
45) L9	AR-1268-5	5.082	4.927	4.972	5.239	5.495	5.143 E5	4.48

(#) = Out of Range