

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR021522\
 Data File : PR053538.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Feb 2022 21:20
 Operator : AJ\MA
 Sample : AR1242ICC100
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12421201

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/16/2022
 Supervised By :Ankita Jodhani 02/16/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 16 03:32:00 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021522CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Feb 16 03:19:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.569	2.815	10282356	6567225	5.117	5.365m
2) SA Decachlor...	9.384	7.963	18922902	15413211	11.012	11.275
Target Compounds						
16) L4 AR-1242-1	4.789	3.899	6143123	4089003	115.014	112.343
17) L4 AR-1242-2	4.810	3.916	8410986	5996755	112.494	112.183
18) L4 AR-1242-3	4.874	4.089	5378308	3185204	114.408	110.694
19) L4 AR-1242-4	4.978	4.181	4306084	3269948	111.483	113.581
20) L4 AR-1242-5	5.758	4.713	4811882	3846136	113.202	106.700

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR021522\
Data File : PR053538.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Feb 2022 21:20
Operator : AJ\MA
Sample : AR1242ICC100
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12421201

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/16/2022
Supervised By :Ankita Jodhani 02/16/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 16 03:32:00 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021522CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Feb 16 03:19:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

