

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR021819CLP.M
 Title : GC EXTRACTABLES
 Last Update : Tue Feb 19 03:57:50 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035571.D 1600 =PR035573.D 800 =PR035572.D
 200 =PR035570.D 100 =PR035569.D

	Compound	400	1600	800	200	100	Avg		%RSD
1)	SA Tetrachloro-m-xylene	1.803	1.695	1.806	1.780	1.863	1.790	E6	3.40
2)	SA Decachlorobiphenyl	3.072	2.933	3.006	3.125	3.301	3.088	E6	4.51
3)	L1 AR-1016-1						0.000		-1.00
4)	L1 AR-1016-2						0.000		-1.00
5)	L1 AR-1016-3						0.000		-1.00
6)	L1 AR-1016-4						0.000		-1.00
7)	L1 AR-1016-5						0.000		-1.00
8)	L2 AR-1221-1						0.000		-1.00
9)	L2 AR-1221-2						0.000		-1.00
10)	L2 AR-1221-3						0.000		-1.00
11)	L3 AR-1232-1						0.000		-1.00
12)	L3 AR-1232-2						0.000		-1.00
13)	L3 AR-1232-3						0.000		-1.00
14)	L3 AR-1232-4						0.000		-1.00
15)	L3 AR-1232-5						0.000		-1.00
16)	L4 AR-1242-1						0.000		-1.00
17)	L4 AR-1242-2						0.000		-1.00
18)	L4 AR-1242-3						0.000		-1.00
19)	L4 AR-1242-4						0.000		-1.00
20)	L4 AR-1242-5						0.000		-1.00
21)	L5 AR-1248-1						0.000		-1.00
22)	L5 AR-1248-2						0.000		-1.00
23)	L5 AR-1248-3						0.000		-1.00
24)	L5 AR-1248-4						0.000		-1.00
25)	L5 AR-1248-5						0.000		-1.00
26)	L6 AR-1254-1						0.000		-1.00
27)	L6 AR-1254-2						0.000		-1.00
28)	L6 AR-1254-3						0.000		-1.00
29)	L6 AR-1254-4						0.000		-1.00
30)	L6 AR-1254-5						0.000		-1.00
31)	L7 AR-1260-1						0.000		-1.00
32)	L7 AR-1260-2						0.000		-1.00
33)	L7 AR-1260-3						0.000		-1.00
34)	L7 AR-1260-4						0.000		-1.00
35)	L7 AR-1260-5						0.000		-1.00
36)	L8 AR-1262-1	3.123	3.013	3.167	3.325	3.056	3.137	E5	3.85
37)	L8 AR-1262-2	5.135	4.966	5.204	5.104	4.838	5.050	E5	2.90
38)	L8 AR-1262-3	3.152	3.070	3.135	3.059	3.109	3.105	E5	1.30
39)	L8 AR-1262-4	2.211	2.142	2.187	2.156	2.071	2.153	E5	2.48
40)	L8 AR-1262-5	1.567	1.540	1.562	1.588	1.647	1.581	E5	2.57
41)	L9 AR-1268-1						0.000		-1.00
42)	L9 AR-1268-2						0.000		-1.00
43)	L9 AR-1268-3						0.000		-1.00
44)	L9 AR-1268-4						0.000		-1.00
45)	L9 AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR035571.D 1600 =PR035573.D 800 =PR035572.D
 200 =PR035570.D 100 =PR035569.D

	Compound	400	1600	800	200	100	Avg		%RSD
1)	SA Tetrachloro-m-xylene	3.260	3.099	3.329	3.205	3.183	3.215	E6	2.67
2)	SA Decachlorobiphenyl	6.789	6.533	6.653	6.868	6.848	6.738	E6	2.11
3)	L1 AR-1016-1						0.000		-1.00
4)	L1 AR-1016-2						0.000		-1.00
5)	L1 AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR021819CLP.M
 Title : GC EXTRACTABLES
 Last Update : Tue Feb 19 03:57:50 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035571.D 1600 =PR035573.D 800 =PR035572.D
 200 =PR035570.D 100 =PR035569.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	5.374	5.113	5.385	5.325	5.421	5.324 E5	2.31
37) L8	AR-1262-2	1.033	0.998	1.042	1.007	1.003	1.017 E6	1.91
38) L8	AR-1262-3	4.140	3.963	4.108	4.144	4.190	4.109 E5	2.11
39) L8	AR-1262-4	7.340	7.104	7.340	7.239	7.278	7.260 E5	1.34
40) L8	AR-1262-5	3.407	3.322	3.373	3.414	3.408	3.385 E5	1.14
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range