

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR021919CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Feb 20 01:01:31 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035635.D 1600 =PR035637.D 800 =PR035636.D
 200 =PR035634.D 100 =PR035633.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	9.130	7.887	8.281	9.258	8.502	8.612	E5	6.70
2)	SA	Decachlorobiphenyl	2.028	1.783	1.904	2.142	2.194	2.010	E6	8.40
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1	6.317	5.021	5.366	6.595	6.838	6.027	E4	13.15
27)	L6	AR-1254-2	1.047	0.851	0.902	1.114	1.118	1.006	E5	12.25
28)	L6	AR-1254-3	1.208	1.014	1.069	1.278	1.282	1.170	E5	10.49
29)	L6	AR-1254-4	9.314	7.798	8.243	9.819	9.673	8.970	E4	10.02
30)	L6	AR-1254-5	1.072	0.910	0.976	1.128	1.129	1.043	E5	9.31
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR035635.D 1600 =PR035637.D 800 =PR035636.D
 200 =PR035634.D 100 =PR035633.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.216	1.947	2.014	2.275	2.091	2.109	E6	6.48
2) SA	Decachlorobiphenyl	6.286	5.880	6.137	6.365	6.275	6.188	E6	3.09
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR021919CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Feb 20 01:01:31 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035635.D 1600 =PR035637.D 800 =PR035636.D
 200 =PR035634.D 100 =PR035633.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	2.309	1.944	2.012	2.363	2.402	2.206 E5	9.60
27) L6	AR-1254-2	2.085	1.696	1.835	2.125	2.136	1.975 E5	10.05
28) L6	AR-1254-3	3.725	3.227	3.355	3.771	3.692	3.554 E5	6.92
29) L6	AR-1254-4	2.687	2.333	2.427	2.776	2.699	2.585 E5	7.45
30) L6	AR-1254-5	3.745	3.325	3.457	3.830	3.654	3.602 E5	5.77
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range