

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR021919\
 Data File : PR035621.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Feb 2019 17:21
 Operator : SM\SJ
 Sample : AR1221ICC100
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 20 01:47:50 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021919CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Feb 20 01:31:01 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.433	3.507	5041867	9996305	5.000	5.000
2)	SA Decachlor...	10.109	8.386	21822061	58263319	10.000	10.000
Target Compounds							
8)	L2 AR-1221-1	4.642	3.716	1349240	3171799	100.000	100.000
9)	L2 AR-1221-2	4.727	3.799	1163501	2404291	100.000	100.000
10)	L2 AR-1221-3	4.804	3.872	3787201	8319676	100.000	100.000
11)	L3 AR-1232-1	4.804	3.872	3787201	8319676	135.859	149.515
12)	L3 AR-1232-2	5.325	4.579	423058	1464128	25.099	18.730 #
13)	L3 AR-1232-3	5.616	0.000	1914214	0	57.313	N.D. #
43)	L9 AR-1268-3	8.964	7.594	706393	1276708	1.947	0.921 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR021919\
Data File : PR035621.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Feb 2019 17:21
Operator : SM\SJ
Sample : AR1221ICC100
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 20 01:47:50 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021919CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Feb 20 01:31:01 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

