



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022024\
Data File : PR065706.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Feb 2024 15:56
Operator : AJ\MA
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 20 20:53:13 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021624CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Feb 17 06:03:42 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(26) AR-1254-1 (L6)

R.T.	Response	Conc
5.79	65771075	401.34
6.03	96891453	405.76
6.42	91655977	406.02
6.73	49058644	391.97
7.18	64177276	413.08

(26) AR-1254-1 #2 (L6)



R.T.	Response	Conc
4.88	113158874	336.31
5.04	98573557	335.12
5.46	157593145	348.57
5.71	86220692	351.64
6.15	133905346	350.58

(+) = Expected Retention Time

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022024\
Data File : PR065706.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Feb 2024 15:56
Operator : AJ\MA
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 20 20:53:13 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021624CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Feb 17 06:03:42 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(26) AR-1254-1 (L6)			
R.T.	Response	Conc	
5.79	65771075	401.34	
6.03	96891453	405.76	
6.42	91655977	406.02	
6.73	54442923	434.99	
7.18	64177276	413.08	

(26) AR-1254-1 #2 (L6)			
R.T.	Response	Conc	
4.88	113158874	336.31	
5.04	98573557	335.12	
5.46	157593145	348.57	
5.71	86220692	351.64	
6.15	133905346	350.58	

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022024\
 Data File : PR065706.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Feb 2024 15:56
 Operator : AJ\MA
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 20 20:53:13 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021624CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Feb 17 06:03:42 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.666	2.929	121.3E6	114.0E6	21.355	18.210
2) SA Decachlor...	9.547	8.193	63488378	97471420	47.039	43.971
Target Compounds						
26) L6 AR-1254-1	5.791	4.882	65771075	113.2E6	401.339	336.307
27) L6 AR-1254-2	6.027	5.041	96891453	98573557	405.764	335.117
28) L6 AR-1254-3	6.418	5.463	91655977	157.6E6	406.021	348.570
29) L6 AR-1254-4	6.728	5.709	54442923	86220692	434.991m	351.643
30) L6 AR-1254-5	7.181	6.155	64177276	133.9E6	413.079	350.583

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022024\
Data File : PR065706.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Feb 2024 15:56
Operator : AJ\MA
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 20 20:53:13 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021624CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Feb 17 06:03:42 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

