

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR022119CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Feb 22 02:34:35 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035684.D 1600 =PR035686.D 800 =PR035685.D
 200 =PR035683.D 100 =PR035682.D

	Compound	400	1600	800	200	100	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.320	1.334	1.263	1.087	1.300	1.261 E6	8.00
2) SA	Decachlorobiphenyl	1.564	1.484	1.513	1.853	1.766	1.636 E6	10.03
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1	1.586	1.543	1.539	1.529	1.825	1.604 E4	7.79
9) L2	AR-1221-2	1.166	1.077	1.098	1.056	1.282	1.136 E4	8.08
10) L2	AR-1221-3	4.038	3.729	3.732	3.615	4.405	3.904 E4	8.23
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

400 =PR035684.D 1600 =PR035686.D 800 =PR035685.D
 200 =PR035683.D 100 =PR035682.D

	Compound	400	1600	800	200	100	Avg	%RSD
1) SA	Tetrachloro-m-xylene	3.190	3.401	3.125	2.617	3.139	3.094 E6	9.34
2) SA	Decachlorobiphenyl	5.606	5.768	5.673	6.226	5.828	5.820 E6	4.16
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR022119CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Feb 22 02:34:35 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035684.D 1600 =PR035686.D 800 =PR035685.D
 200 =PR035683.D 100 =PR035682.D

	Compound	400	1600	800	200	100	Avg	%RSD

6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1	4.834	4.702	4.580	4.032	4.872	4.604 E4	7.39
9) L2	AR-1221-2	3.473	3.314	3.273	3.080	3.899	3.408 E4	9.04
10) L2	AR-1221-3	1.220	1.189	1.161	1.095	1.324	1.198 E5	7.06
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

 (#) = Out of Range