

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR022119CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Feb 22 01:40:23 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035694.D 1600 =PR035696.D 800 =PR035695.D
 200 =PR035693.D 100 =PR035692.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.325	1.210	1.015	1.100	1.253	1.181	E6	10.45
2) SA	Decachlorobiphenyl	1.581	1.449	1.570	1.832	1.759	1.638	E6	9.45
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1	4.483	3.808	3.338	3.930	4.682	4.048	E4	13.34
17) L4	AR-1242-2	7.370	6.433	5.519	6.435	7.404	6.632	E4	11.82
18) L4	AR-1242-3	4.472	3.746	3.365	4.034	4.788	4.081	E4	13.87
19) L4	AR-1242-4	3.654	3.122	2.764	3.255	3.861	3.331	E4	13.05
20) L4	AR-1242-5	4.587	3.877	3.547	4.124	4.923	4.212	E4	13.04
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR035694.D 1600 =PR035696.D 800 =PR035695.D
 200 =PR035693.D 100 =PR035692.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	4.075	3.810	3.095	3.305	3.731	3.603	E6	11.00
2) SA	Decachlorobiphenyl	5.769	5.995	6.271	6.427	6.050	6.102	E6	4.17
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR022119CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Feb 22 01:40:23 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035694.D 1600 =PR035696.D 800 =PR035695.D
 200 =PR035693.D 100 =PR035692.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1	1.484	1.326	1.095	1.252	1.446	1.321 E5	11.85
17) L4	AR-1242-2	2.325	2.100	1.760	1.962	2.228	2.075 E5	10.73
18) L4	AR-1242-3	1.204	1.075	0.911	1.025	1.152	1.073 E5	10.63
19) L4	AR-1242-4	1.167	1.014	0.877	1.012	1.165	1.047 E5	11.64
20) L4	AR-1242-5	1.724	1.523	1.333	1.525	1.759	1.573 E5	11.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range