

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR022119CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Feb 22 01:24:24 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035714.D 1600 =PR035716.D 800 =PR035715.D
 200 =PR035713.D 100 =PR035712.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.319	1.128	0.958	1.314	1.278	1.199	E6	12.98
2) SA	Decachlorobiphenyl	3.130	3.264	3.362	3.377	3.544	3.335	E6	4.58
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1	3.798	3.693	3.595	3.966	4.093	3.829	E5	5.27
42) L9	AR-1268-2	3.359	3.302	3.214	3.540	3.655	3.414	E5	5.27
43) L9	AR-1268-3	2.908	2.855	2.781	3.046	3.120	2.942	E5	4.72
44) L9	AR-1268-4	1.088	1.080	1.119	1.186	1.195	1.133	E5	4.74
45) L9	AR-1268-5	8.299	9.027	9.233	8.566	8.760	8.777	E5	4.20

Signal #2 Calibration Files

400 =PR035714.D 1600 =PR035716.D 800 =PR035715.D
 200 =PR035713.D 100 =PR035712.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	4.088	3.547	3.000	4.043	3.715	3.679	E6	12.00
2) SA	Decachlorobiphenyl	1.305	1.382	1.429	1.349	1.330	1.359	E7	3.56
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR022119CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Feb 22 01:24:24 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035714.D 1600 =PR035716.D 800 =PR035715.D
 200 =PR035713.D 100 =PR035712.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	1.995	1.877	1.873	2.032	1.938	1.943 E6	3.63
42) L9	AR-1268-2	1.712	1.659	1.640	1.737	1.693	1.688 E6	2.32
43) L9	AR-1268-3	1.519	1.471	1.455	1.542	1.482	1.494 E6	2.38
44) L9	AR-1268-4	4.991	5.301	5.350	5.095	5.164	5.180 E5	2.84
45) L9	AR-1268-5	3.872	3.949	4.180	3.961	3.863	3.965 E6	3.23

(#) = Out of Range