

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022123\
 Data File : PR059614.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Feb 2023 15:16
 Operator : YP\AJ
 Sample : 01601-06
 Misc :
 ALS Vial : 17 Sample Multiplier: 1



Instrument :
 ECD_R
 ClientSampleId :
 EXHD5

Manual Integrations APPROVED

Reviewed By : Yogesh Patel 02/22/2023
 Supervised By : Ankita Jodhani 02/22/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 02:09:51 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 27 08:44:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(2) Decachlorobiphenyl (SA)

9.665min 28.150 ng/ml

response 58876452

(2) Decachlorobiphenyl #2 (SA)

8.142min 22.806 ng/ml

response 72766789

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022123\
 Data File : PR059614.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Feb 2023 15:16
 Operator : YP\AJ
 Sample : 01601-06
 Misc :
 ALS Vial : 17 Sample Multiplier: 1



Instrument :
 ECD_R
 ClientSampleId :
 EXHD5

Manual Integrations APPROVED

Reviewed By : Yogesh Patel 02/22/2023
 Supervised By : Ankita Jodhani 02/22/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 02:09:51 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 27 08:44:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(2) Decachlorobiphenyl (SA)

9.664min 31.875 ng/ml m

response 66666362

(2) Decachlorobiphenyl #2 (SA)

8.141min 27.478 ng/ml m

response 87674077

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022123\
 Data File : PR059614.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Feb 2023 15:16
 Operator : YP\AJ
 Sample : 01601-06
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :

ECD_R

ClientSampleId :



EXHD5

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 02/22/2023
 Supervised By :Ankita Jodhani 02/22/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 02:09:51 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 27 08:44:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(21) AR-1248-1 (L5)

R.T.	Response	Conc
4.92	6296218	111.91
5.22	6496929	89.12
5.43	5516056	64.23
5.87	10425841	109.56
5.91	11626047	126.08

(21) AR-1248-1 #2 (L5)

R.T.	Response	Conc
4.02	7105474	81.83
4.26	7138050	60.88
4.30	7033169	58.57
4.48	5872531	39.88
4.89	24322301	164.69

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022123\
 Data File : PR059614.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Feb 2023 15:16
 Operator : YP\AJ
 Sample : 01601-06
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :

ECD_R

ClientSampleId :



EXHD5

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 02/22/2023
 Supervised By :Ankita Jodhani 02/22/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 02:09:51 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 27 08:44:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(21) AR-1248-1 (L5)

R.T.	Response	Conc
4.92	6296218	111.91
5.22	6496929	89.12
5.43	5516056	64.23
5.87	10425841	109.56
5.91	11626047	126.08

(21) AR-1248-1 #2 (L5)

R.T.	Response	Conc
4.02	5318667	61.25
4.26	7138050	60.88
4.30	7033169	58.57
4.48	6041287	41.02
4.89	24322301	164.69

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022123\
 Data File : PR059614.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Feb 2023 15:16
 Operator : YP\AJ
 Sample : 01601-06
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :

ECD_R

ClientSampleId :



EXHD5

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 02/22/2023
 Supervised By :Ankita Jodhani 02/22/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 02:09:51 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 27 08:44:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)

R.T.	Response	Conc
6.66	9505368	81.18
6.94	19906269	143.45
7.33	10566805	102.33
7.57	33417002	285.28
7.92	15936963	70.13

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
5.57	15088833	70.62
5.77	25396427	98.76
5.93	12668202	53.76
6.43	12555669	64.98
6.70	29133002	64.98

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022123\
 Data File : PR059614.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Feb 2023 15:16
 Operator : YP\AJ
 Sample : 01601-06
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :

ECD_R

ClientSampleId :



EXHD5

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 02/22/2023
 Supervised By :Ankita Jodhani 02/22/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 02:09:51 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 27 08:44:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)

R.T.	Response	Conc
6.66	9505368	81.18
6.94	19906269	143.45
7.33	10566805	102.33
7.57	32335278	276.04
7.92	15936963	70.13

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
5.57	15088833	70.62
5.77	25396427	98.76
5.93	10844172	46.02
6.43	12555669	64.98
6.70	34221744	76.33