

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022219\
 Data File : PR035737.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Feb 2019 11:29
 Operator : SM\SJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1254326

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 22:47:54 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022119CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 22 02:53:40 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds							
20)	L4	AR-1242-5	6.472	5.360	635416	617807	15.087 3.927 #
24)	L5	AR-1248-4	6.472	0.000	635416	0	10.287 N.D. #
25)	L5	AR-1248-5	6.472	5.360	635416	617807	10.556 3.254 #
26)	L6	AR-1254-1	6.436	5.360	1028463	617807	16.528 2.085 #
27)	L6	AR-1254-2	6.632	0.000	1147510	0	11.357 N.D. #
28)	L6	AR-1254-3	7.032	5.860	519054	5778700	4.610 12.737 #
29)	L6	AR-1254-4	7.287	0.000	120431	0	1.425 N.D. #
32)	L7	AR-1260-2	7.437	6.179	49728	9484566	0.282 13.247 #
34)	L7	AR-1260-4	8.014	0.000	122529	0	0.795 N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022219\
Data File : PR035737.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Feb 2019 11:29
Operator : SM\SJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1254326

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 22:47:54 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022119CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 22 02:53:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#20 AR-1242-5


R.T.: 6.472 min
Delta R.T.: -0.027 min
Response: 635416
Conc: 15.09 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254326




#20 AR-1242-5

R.T.: 5.360 min
Delta R.T.: 0.025 min
Response: 617807
Conc: 3.93 ng/ml




#24 AR-1248-4

R.T.: 6.472 min
Delta R.T.: 0.012 min
Response: 635416
Conc: 10.29 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T.: 4.991 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.472 min
Delta R.T.: -0.027 min
Response: 635416
Conc: 10.56 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254326




#25 AR-1248-5

R.T.: 5.360 min
Delta R.T.: -0.015 min
Response: 617807
Conc: 3.25 ng/ml




#26 AR-1254-1

R.T.: 6.436 min
Delta R.T.: 0.003 min
Response: 1028463
Conc: 16.53 ng/ml



#26 AR-1254-1


R.T.: 5.360 min
Delta R.T.: 0.025 min
Response: 617807
Conc: 2.08 ng/ml



#27 AR-1254-2


R.T.: 6.632 min
Delta R.T.: -0.017 min
Response: 1147510
Conc: 11.36 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254326




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T. : 5.479 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 7.032 min
Delta R.T.: 0.018 min
Response: 519054
Conc: 4.61 ng/ml



#28 AR-1254-3


R.T.: 5.860 min
Delta R.T.: -0.015 min
Response: 5778700
Conc: 12.74 ng/ml



#29 AR-1254-4


R.T.: 7.287 min
Delta R.T.: -0.012 min
Response: 120431
Conc: 1.43 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254326




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.101 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 7.437 min
Delta R.T.: 0.005 min
Response: 49728
Conc: 0.28 ng/ml



#32 AR-1260-2


R.T.: 6.179 min
Delta R.T.: -0.010 min
Response: 9484566
Conc: 13.25 ng/ml



#34 AR-1260-4

R.T.: 8.014 min
Delta R.T.: -0.001 min
Response: 122529
Conc: 0.80 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254326



#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 6.802 min
Response: 0
Conc: N.D.