Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022219\
Data File : PR035769.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Feb 2019 22:32
Operator : SM\SJ
Sample : AR1254CCC400
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1254329

Manual Integrations APPROVED
Sohil
2/25/2019 8:50:46 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 22:59:38 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022119CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 22 02:53:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.421	3.495	23890010	72593402	18.981	21.772
2)	SA Decachlor...	10.091	8.373	53382382	208.6E6	29.989	35.278
	Target Compounds						
26)	L6 AR-1254-1	6.431	5.333	24372613	115.0E6	391.693	388.078
27)	L6 AR-1254-2	6.648	5.478	38584484	102.3E6	381.868	384.633
28)	L6 AR-1254-3	7.014	5.873	42436013	171.1E6	376.863	377.207
29)	L6 AR-1254-4	7.298	6.098	31129903	119.3E6	368.368	374.555
30)	L6 AR-1254-5	7.714	6.508	35716315	162.0E6	372.020m	371.783

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022219\
Data File : PR035769.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Feb 2019 22:32
Operator : SM\SJ
Sample : AR1254CCC400
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1254329

Manual Integrations APPROVED
Sohil
2/25/2019 8:50:46 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 22:59:38 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022119CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 22 02:53:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm