

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022219\
Data File : PR035741.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Feb 2019 14:50
Operator : SM\SJ
Sample : K1455-01 10X
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
BE801

Manual Integrations
APPROVED

Sohil
2/25/2019 8:50:40 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 22:49:26 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022119CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 22 02:53:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

4.412min 17.071 ng/ml

response 21485799

(1) Tetrachloro-m-xylene #2 (SA)

3.498min 0.475 ng/ml

response 1583908

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022219\
Data File : PR035741.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Feb 2019 14:50
Operator : SM\SJ
Sample : K1455-01 10X
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
BE801

Manual Integrations
APPROVED

Sohil
2/25/2019 8:50:40 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 22:49:26 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022119CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 22 02:53:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

4.412min 17.071 ng/ml

response 21485799

(1) Tetrachloro-m-xylene #2 (SA)

3.497min 0.718 ng/ml m

response 2393323

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022219\
 Data File : PR035741.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Feb 2019 14:50
 Operator : SM\SJ
 Sample : K1455-01 10X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 BE801

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 22:49:26 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022119CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 22 02:53:40 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Manual Integrations
 APPROVED

Sohil
 2/25/2019 8:50:40 AM

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	4.412	3.497	21485799	2393323	17.071	0.718m#
2) SA Decachlor...	10.101	8.379	7368064	23820808	4.139	4.029

Target Compounds

3) L1 AR-1016-1	5.592	4.554	587.9E6	1731.2E6	7837.335	7336.266
4) L1 AR-1016-2	5.614	4.571	1036.7E6	2784.2E6	8455.000	7320.524
5) L1 AR-1016-3	5.678	4.743	335.8E6	1290.9E6	4604.243	6731.531 #
6) L1 AR-1016-4	5.774	4.784	430.2E6	1091.7E6	7139.893	7444.982
7) L1 AR-1016-5	6.065	4.992	468.3E6	1413.6E6	7533.452	6851.423
31) L7 AR-1260-1	7.182	6.003	758.6E6	2597.8E6	5357.696	5201.542
32) L7 AR-1260-2	7.437	6.189	1235.4E6	3470.2E6	7008.224	4846.948 #
33) L7 AR-1260-3	7.722	6.339	1344.0E6	2772.5E6	5847.491	4573.038
34) L7 AR-1260-4	8.021	6.804	813.2E6	2292.4E6	5277.215	4466.672
35) L7 AR-1260-5	8.337	7.046	1658.2E6	6870.0E6	5115.276	4679.057

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.