

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022223\
 Data File : PR059682.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Feb 2023 12:39
 Operator : YP\AJ
 Sample : 01631-11
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 EXHJ1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/23/2023
 Supervised By :Ankita Jodhani 02/23/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 20:33:03 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 27 08:44:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.690	2.906	44094687	61353124	16.255	15.430
2) SA Decachlor...	9.663	8.141	51133815	73377916	24.448	22.997
Target Compounds						
26) L6 AR-1254-1	5.841	4.847	2396537	5865051	25.371	26.035
27) L6 AR-1254-2	6.079	5.006	4869180	6546481	33.352m	34.037
28) L6 AR-1254-3	6.475	5.424	7691322	10935664	49.842m	34.811 #
29) L6 AR-1254-4	6.788	5.671	7767635	11667689	67.160m	60.232
30) L6 AR-1254-5	7.246	6.114	11366704	30401803	91.775	112.792

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022223\
Data File : PR059682.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Feb 2023 12:39
Operator : YP\AJ
Sample : 01631-11
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
EXHJ1

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/23/2023
Supervised By :Ankita Jodhani 02/23/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 20:33:03 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 27 08:44:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

