

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022423\
 Data File : PR059772.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Feb 2023 02:53
 Operator : YP\AJ
 Sample : 01627-08
 Misc : AR1221 LOQ 50 PPB
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 LOQ-WATER-02-QT1-2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 08:06:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 24 12:41:32 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.689	2.906	55661541	79964217	18.630	18.474
2) SA Decachlor...	9.658	8.137	51932328	74258696	22.641	21.049
Target Compounds						
8) L2 AR-1221-1	3.913	3.129	1819506	3029584	59.338	79.880 #
9) L2 AR-1221-2	4.003	3.213	1920444	2410524	85.319	85.440
10) L2 AR-1221-3	4.082	3.291	4333904	5279479	64.600	59.724

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022423\
Data File : PR059772.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Feb 2023 02:53
Operator : YP\AJ
Sample : 01627-08
Misc : AR1221 LOQ 50 PPB
ALS Vial : 49 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
LOQ-WATER-02-QT1-2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 08:06:58 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022423.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 24 12:41:32 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

