

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022423\
 Data File : PR059780.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Feb 2023 04:50
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 08:09:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 24 12:41:32 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	3.690	2.905	62452142	89268416	20.903	20.624
2) SA Decachlor...	9.659	8.136	57070094	82329015	24.881	23.337

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022423\
Data File : PR059780.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Feb 2023 04:50
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 08:09:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022423.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 24 12:41:32 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.690 min
Delta R.T.: -0.001 min
Response: 62452142
Conc: 20.90 ng/ml

Instrument :
ECD_R
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.905 min
Delta R.T.: -0.001 min
Response: 89268416
Conc: 20.62 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.659 min
Delta R.T.: -0.007 min
Response: 57070094
Conc: 24.88 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.136 min
Delta R.T.: -0.005 min
Response: 82329015
Conc: 23.34 ng/ml