

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022620\
 Data File : PR044799.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Feb 2020 10:38
 Operator : AJ\MA
 Sample : L1645-08
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 PCB-GPC-BLANK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 26 13:46:03 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022420CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 25 08:20:41 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.674	3.968	111754	199976	0.014	0.013
2)	SA Decachlor...	10.578	9.038	177812	139682	0.024	0.011 #
Target Compounds							
3)	L1 AR-1016-1	0.000	5.047	0	261344	N.D.	0.683 #
6)	L1 AR-1016-4	5.988f	5.279	65749	474648	0.304	2.027 #
7)	L1 AR-1016-5	6.401f	5.502	941983	386867	4.603	1.212 #
8)	L2 AR-1221-1	4.882	4.162	939341	83732	9.361	0.459 #
9)	L2 AR-1221-2	4.991	4.245	2172865	67832	32.069	0.527 #
10)	L2 AR-1221-3	5.112f	4.313	470381	111180	2.099	0.268 #
11)	L3 AR-1232-1	5.112f	4.313	470381	111180	2.537	0.325 #
12)	L3 AR-1232-2	0.000	5.059	0	322658	N.D.	1.247 #
14)	L3 AR-1232-4	5.988f	5.323	65749	298622	0.630	2.473 #
15)	L3 AR-1232-5	6.134	5.502	200821	386867	2.691	2.825
16)	L4 AR-1242-1	0.000	5.047	0	261344	N.D.	0.784 #
19)	L4 AR-1242-4	5.988f	5.323	65749	298622	0.333	1.223 #
20)	L4 AR-1242-5	6.856f	5.850	1004828	236972	4.597	0.620 #
21)	L5 AR-1248-1	0.000	5.047	0	261344	N.D.	0.968 #
22)	L5 AR-1248-2	6.134	5.279	200821	474648	0.668	1.351 #
23)	L5 AR-1248-3	6.401f	5.323	941983	298622	2.832	0.819 #
24)	L5 AR-1248-4	6.677f	5.502	1061521	386867	2.726	0.815 #
25)	L5 AR-1248-5	0.000	5.891	0	642126	N.D.	1.237 #
26)	L6 AR-1254-1	6.677f	5.850	1061521	236972	2.754	0.297 #
27)	L6 AR-1254-2	6.927	5.990	1963884	664263	3.312	0.935 #
28)	L6 AR-1254-3	7.333	6.406	6237481	575373	9.967	0.488 #
29)	L6 AR-1254-4	7.582	6.622	3700630	952747	7.836	1.090 #
30)	L6 AR-1254-5	8.067f	7.068	1281581	2238152	2.610	2.092
31)	L7 AR-1260-1	0.000	6.536	0	1178662	N.D.	1.760 #
32)	L7 AR-1260-2	7.714	6.712	1202704	2995736	2.550	2.761
33)	L7 AR-1260-3	8.067	6.898	1281581	9645075	3.457	10.996 #
34)	L7 AR-1260-4	8.361f	7.326	1012887	7225959	2.418	9.464 #
35)	L7 AR-1260-5	8.711f	7.589	623851	3443057	0.689	1.557 #
36)	L8 AR-1262-1	8.120	7.068	738991	2238152	3.071	3.829
37)	L8 AR-1262-2	8.711f	7.589	623851	3443057	0.549	1.313 #
38)	L8 AR-1262-3	8.938f	7.908f	1677349	10774419	2.247	10.128 #
39)	L8 AR-1262-4	9.093	7.908	1847990	10774419	3.208	6.173 #
40)	L8 AR-1262-5	9.788	8.446	774064	1005843	1.932	1.275 #
41)	L9 AR-1268-1	8.938f	7.908f	1677349	10774419	1.237	3.520 #
42)	L9 AR-1268-2	9.093	7.908	1847990	10774419	1.446	4.131 #
43)	L9 AR-1268-3	9.280f	8.121	4528909	4537738	4.246	2.035 #
44)	L9 AR-1268-4	9.788	8.446	774064	1005843	1.714	1.130 #
45)	L9 AR-1268-5	10.230	8.757	726311	468837	0.202	0.075 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022620\
Data File : PR044799.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Feb 2020 10:38
Operator : AJ\MA
Sample : L1645-08
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 26 13:46:03 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022420CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 25 08:20:41 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022620\
Data File : PR044799.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Feb 2020 10:38
Operator : AJ\MA
Sample : L1645-08
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 26 13:46:03 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022420CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 25 08:20:41 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#3 AR-1016-1


R.T.: 0.000 min
Exp R.T.: 5.865 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#3 AR-1016-1

R.T.: 5.047 min
Delta R.T.: 0.006 min
Response: 261344
Conc: 0.68 ng/ml




#4 AR-1016-2

R.T.: 5.956 min
Delta R.T.: 0.067 min
Response: -313352
Conc: N.D.



#4 AR-1016-2


R.T.: 5.059 min
Delta R.T.: -0.003 min
Response: 322658
Conc: 0.59 ng/ml



#5 AR-1016-3


R.T.: 5.956 min
Delta R.T.: 0.005 min
Response: -313352
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#5 AR-1016-3

R.T.: 5.224 min
Delta R.T.: -0.016 min
Response: 103275
Conc: 0.34 ng/ml




#6 AR-1016-4

R.T.: 5.988 min
Delta R.T.: -0.063 min
Response: 65749
Conc: 0.30 ng/ml



#6 AR-1016-4


R.T.: 5.279 min
Delta R.T.: 0.000 min
Response: 474648
Conc: 2.03 ng/ml



#7 AR-1016-5


R.T.: 6.401 min
Delta R.T.: 0.057 min
Response: 941983
Conc: 4.60 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#7 AR-1016-5

R.T.: 5.502 min
Delta R.T.: 0.006 min
Response: 386867
Conc: 1.21 ng/ml




#8 AR-1221-1

R.T.: 4.882 min
Delta R.T.: -0.015 min
Response: 939341
Conc: 9.36 ng/ml



#8 AR-1221-1


R.T.: 4.162 min
Delta R.T.: -0.001 min
Response: 83732
Conc: 0.46 ng/ml



#9 AR-1221-2


R.T.: 4.991 min
Delta R.T.: 0.005 min
Response: 2172865
Conc: 32.07 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#9 AR-1221-2

R.T.: 4.245 min
Delta R.T.: -0.005 min
Response: 67832
Conc: 0.53 ng/ml




#10 AR-1221-3

R.T.: 5.112 min
Delta R.T.: 0.049 min
Response: 470381
Conc: 2.10 ng/ml



#10 AR-1221-3


R.T.: 4.313 min
Delta R.T.: -0.013 min
Response: 111180
Conc: 0.27 ng/ml



#11 AR-1232-1


R.T.: 5.112 min
Delta R.T.: 0.049 min
Response: 470381
Conc: 2.54 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#11 AR-1232-1

R.T.: 4.313 min
Delta R.T.: -0.013 min
Response: 111180
Conc: 0.32 ng/ml




#12 AR-1232-2

R.T.: 0.000 min
Exp R.T. : 5.597 min
Response: 0
Conc: N.D.



#12 AR-1232-2


R.T.: 5.059 min
Delta R.T.: -0.002 min
Response: 322658
Conc: 1.25 ng/ml



#13 AR-1232-3


R.T.: 5.956 min
Delta R.T.: 0.068 min
Response: -313352
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#13 AR-1232-3

R.T.: 5.224 min
Delta R.T.: -0.016 min
Response: 103275
Conc: 0.73 ng/ml




#14 AR-1232-4

R.T.: 5.988 min
Delta R.T.: -0.062 min
Response: 65749
Conc: 0.63 ng/ml



#14 AR-1232-4


R.T.: 5.323 min
Delta R.T.: 0.001 min
Response: 298622
Conc: 2.47 ng/ml



#15 AR-1232-5


R.T.: 6.134 min
Delta R.T.: -0.002 min
Response: 200821
Conc: 2.69 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#15 AR-1232-5

R.T.: 5.502 min
Delta R.T.: 0.007 min
Response: 386867
Conc: 2.83 ng/ml




#16 AR-1242-1

R.T.: 0.000 min
Exp R.T. : 5.864 min
Response: 0
Conc: N.D.



#16 AR-1242-1


R.T.: 5.047 min
Delta R.T.: 0.006 min
Response: 261344
Conc: 0.78 ng/ml



#17 AR-1242-2


R.T.: 5.956 min
Delta R.T.: 0.069 min
Response: -313352
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#17 AR-1242-2

R.T.: 5.059 min
Delta R.T.: -0.002 min
Response: 322658
Conc: 0.70 ng/ml




#18 AR-1242-3

R.T.: 5.956 min
Delta R.T.: 0.007 min
Response: -313352
Conc: N.D.



#18 AR-1242-3


R.T.: 5.224 min
Delta R.T.: -0.016 min
Response: 103275
Conc: 0.40 ng/ml



#19 AR-1242-4


R.T.: 5.988 min
Delta R.T.: -0.062 min
Response: 65749
Conc: 0.33 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#19 AR-1242-4

R.T.: 5.323 min
Delta R.T.: 0.000 min
Response: 298622
Conc: 1.22 ng/ml




#20 AR-1242-5

R.T.: 6.856 min
Delta R.T.: 0.070 min
Response: 1004828
Conc: 4.60 ng/ml



#20 AR-1242-5


R.T.: 5.850 min
Delta R.T.: 0.003 min
Response: 236972
Conc: 0.62 ng/ml



#21 AR-1248-1


R.T.: 0.000 min
Exp R.T.: 5.864 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#21 AR-1248-1

R.T.: 5.047 min
Delta R.T.: 0.005 min
Response: 261344
Conc: 0.97 ng/ml




#22 AR-1248-2

R.T.: 6.134 min
Delta R.T.: -0.002 min
Response: 200821
Conc: 0.67 ng/ml



#22 AR-1248-2


R.T.: 5.279 min
Delta R.T.: 0.000 min
Response: 474648
Conc: 1.35 ng/ml



#23 AR-1248-3


R.T.: 6.401 min
Delta R.T.: 0.058 min
Response: 941983
Conc: 2.83 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#23 AR-1248-3

R.T.: 5.323 min
Delta R.T.: 0.000 min
Response: 298622
Conc: 0.82 ng/ml




#24 AR-1248-4

R.T.: 6.677 min
Delta R.T.: -0.070 min
Response: 1061521
Conc: 2.73 ng/ml



#24 AR-1248-4


R.T.: 5.502 min
Delta R.T.: 0.008 min
Response: 386867
Conc: 0.81 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 6.785 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#25 AR-1248-5

R.T.: 5.891 min
Delta R.T.: 0.003 min
Response: 642126
Conc: 1.24 ng/ml




#26 AR-1254-1

R.T.: 6.677 min
Delta R.T.: -0.042 min
Response: 1061521
Conc: 2.75 ng/ml



#26 AR-1254-1


R.T.: 5.850 min
Delta R.T.: 0.004 min
Response: 236972
Conc: 0.30 ng/ml



#27 AR-1254-2


R.T.: 6.927 min
Delta R.T.: -0.008 min
Response: 1963884
Conc: 3.31 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#27 AR-1254-2

R.T.: 5.990 min
Delta R.T.: -0.002 min
Response: 664263
Conc: 0.93 ng/ml




#28 AR-1254-3

R.T.: 7.333 min
Delta R.T.: 0.028 min
Response: 6237481
Conc: 9.97 ng/ml



#28 AR-1254-3


R.T.: 6.406 min
Delta R.T.: 0.008 min
Response: 575373
Conc: 0.49 ng/ml



#29 AR-1254-4


R.T.: 7.582 min
Delta R.T.: -0.007 min
Response: 3700630
Conc: 7.84 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#29 AR-1254-4

R.T.: 6.622 min
Delta R.T.: -0.003 min
Response: 952747
Conc: 1.09 ng/ml




#30 AR-1254-5

R.T.: 8.067 min
Delta R.T.: 0.059 min
Response: 1281581
Conc: 2.61 ng/ml



#30 AR-1254-5


R.T.: 7.068 min
Delta R.T.: 0.025 min
Response: 2238152
Conc: 2.09 ng/ml



#31 AR-1260-1


R.T.: 0.000 min
Exp R.T.: 7.468 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#31 AR-1260-1

R.T.: 6.536 min
Delta R.T.: 0.008 min
Response: 1178662
Conc: 1.76 ng/ml




#32 AR-1260-2

R.T.: 7.714 min
Delta R.T.: -0.008 min
Response: 1202704
Conc: 2.55 ng/ml



#32 AR-1260-2


R.T.: 6.712 min
Delta R.T.: -0.004 min
Response: 2995736
Conc: 2.76 ng/ml



#33 AR-1260-3


R.T.: 8.067 min
Delta R.T.: -0.015 min
Response: 1281581
Conc: 3.46 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#33 AR-1260-3

R.T.: 6.898 min
Delta R.T.: 0.028 min
Response: 9645075
Conc: 11.00 ng/ml




#34 AR-1260-4

R.T.: 8.361 min
Delta R.T.: 0.041 min
Response: 1012887
Conc: 2.42 ng/ml



#34 AR-1260-4


R.T.: 7.326 min
Delta R.T.: -0.015 min
Response: 7225959
Conc: 9.46 ng/ml



#35 AR-1260-5


R.T.: 8.711 min
Delta R.T.: 0.056 min
Response: 623851
Conc: 0.69 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#35 AR-1260-5

R.T.: 7.589 min
Delta R.T.: 0.009 min
Response: 3443057
Conc: 1.56 ng/ml




#36 AR-1262-1

R.T.: 8.120 min
Delta R.T.: -0.020 min
Response: 738991
Conc: 3.07 ng/ml



#36 AR-1262-1


R.T.: 7.068 min
Delta R.T.: 0.026 min
Response: 2238152
Conc: 3.83 ng/ml



#37 AR-1262-2


R.T.: 8.711 min
Delta R.T.: 0.058 min
Response: 623851
Conc: 0.55 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#37 AR-1262-2

R.T.: 7.589 min
Delta R.T.: 0.009 min
Response: 3443057
Conc: 1.31 ng/ml




#38 AR-1262-3

R.T.: 8.938 min
Delta R.T.: -0.043 min
Response: 1677349
Conc: 2.25 ng/ml



#38 AR-1262-3


R.T.: 7.908 min
Delta R.T.: 0.045 min
Response: 10774419
Conc: 10.13 ng/ml



#39 AR-1262-4


R.T.: 9.093 min
Delta R.T.: 0.017 min
Response: 1847990
Conc: 3.21 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#39 AR-1262-4

R.T.: 7.908 min
Delta R.T.: -0.021 min
Response: 10774419
Conc: 6.17 ng/ml




#40 AR-1262-5

R.T.: 9.788 min
Delta R.T.: 0.012 min
Response: 774064
Conc: 1.93 ng/ml



#40 AR-1262-5


R.T.: 8.446 min
Delta R.T.: 0.005 min
Response: 1005843
Conc: 1.27 ng/ml



#41 AR-1268-1


R.T.: 8.938 min
Delta R.T.: -0.042 min
Response: 1677349
Conc: 1.24 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#41 AR-1268-1

R.T.: 7.908 min
Delta R.T.: 0.045 min
Response: 10774419
Conc: 3.52 ng/ml




#42 AR-1268-2

R.T.: 9.093 min
Delta R.T.: 0.014 min
Response: 1847990
Conc: 1.45 ng/ml



#42 AR-1268-2


R.T.: 7.908 min
Delta R.T.: -0.022 min
Response: 10774419
Conc: 4.13 ng/ml



#43 AR-1268-3


R.T.: 9.280 min
Delta R.T.: -0.040 min
Response: 4528909
Conc: 4.25 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK




#43 AR-1268-3

R.T.: 8.121 min
Delta R.T.: -0.015 min
Response: 4537738
Conc: 2.03 ng/ml




#44 AR-1268-4

R.T.: 9.788 min
Delta R.T.: 0.013 min
Response: 774064
Conc: 1.71 ng/ml



#44 AR-1268-4


R.T.: 8.446 min
Delta R.T.: 0.007 min
Response: 1005843
Conc: 1.13 ng/ml



#45 AR-1268-5

R.T.: 10.230 min
Delta R.T.: 0.015 min
Response: 726311
Conc: 0.20 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GPC-BLANK



#45 AR-1268-5

R.T.: 8.757 min
Delta R.T.: 0.006 min
Response: 468837
Conc: 0.07 ng/ml