

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030121\
 Data File : PR049472.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Mar 2021 21:42
 Operator : AJ\MA
 Sample : AR1254ICC400
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 02 03:44:17 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR030121CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 02 03:43:00 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.664	3.849	55777426	68708990	20.000	20.000
2) SA Decachlor...	10.545	8.900	110.4E6	365.4E6	40.000	40.000
Target Compounds						
26) L6 AR-1254-1	6.696	5.735	44165630	48262314	400.000	400.000
27) L6 AR-1254-2	6.913	5.882	68243031	53533342	400.000	400.000
28) L6 AR-1254-3	7.283	6.287	70859469	103.8E6	400.000	400.000
29) L6 AR-1254-4	7.568	6.516	58329168	165.4E6	400.000	400.000
30) L6 AR-1254-5	7.986	6.934	56915769	196.2E6	400.000	400.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030121\
Data File : PR049472.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Mar 2021 21:42
Operator : AJ\MA
Sample : AR1254ICC400
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 02 03:44:17 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR030121CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 02 03:43:00 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#26 AR-1254-1


R.T.: 6.696 min
Delta R.T.: 0.000 min
Response: 44165630
Conc: 400.00 ng/ml

Instrument :
ECD_R
ClientSampleId :




#26 AR-1254-1

R.T.: 5.735 min
Delta R.T.: 0.000 min
Response: 48262314
Conc: 400.00 ng/ml




#27 AR-1254-2

R.T.: 6.913 min
Delta R.T.: 0.000 min
Response: 68243031
Conc: 400.00 ng/ml



#27 AR-1254-2


R.T.: 5.882 min
Delta R.T.: 0.000 min
Response: 53533342
Conc: 400.00 ng/ml



#28 AR-1254-3


R.T.: 7.283 min
Delta R.T.: 0.000 min
Response: 70859469
Conc: 400.00 ng/ml

Instrument :
ECD_R
ClientSampleId :




#28 AR-1254-3

R.T.: 6.287 min
Delta R.T.: 0.000 min
Response: 103777108
Conc: 400.00 ng/ml




#29 AR-1254-4

R.T.: 7.568 min
Delta R.T.: 0.000 min
Response: 58329168
Conc: 400.00 ng/ml



#29 AR-1254-4


R.T.: 6.516 min
Delta R.T.: 0.000 min
Response: 165442705
Conc: 400.00 ng/ml



#30 AR-1254-5

R.T.: 7.986 min
Delta R.T.: 0.000 min
Response: 56915769
Conc: 400.00 ng/ml

Instrument :
ECD_R
ClientSampleId :



#30 AR-1254-5

R.T.: 6.934 min
Delta R.T.: 0.000 min
Response: 196199068
Conc: 400.00 ng/ml