

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030122\
 Data File : PR053702.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Mar 2022 07:23
 Operator : AJ\MA
 Sample : N1693-19
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 BGZQ2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/01/2022
 Supervised By :Ankita Jodhani 03/04/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 09:35:39 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021522CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Feb 16 05:21:54 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.566	2.813	21833833	13842495	11.146	11.511
2) SA Decachlor...	9.374	7.956	32859002	26142688	20.170	19.485
Target Compounds						
26) L6 AR-1254-1	5.685	4.711	254743	388989	3.566	5.185m#
27) L6 AR-1254-2	5.917	4.864	1420653	435512	12.612	6.625m#
28) L6 AR-1254-3	6.311	5.276	1090852	583119	9.291	5.656m#
29) L6 AR-1254-4	6.618	5.521	1988915	769858	21.914m	11.695 #
30) L6 AR-1254-5	7.070	5.956	3642710	2614804	37.575	27.785 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030122\
Data File : PR053702.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Mar 2022 07:23
Operator : AJ\MA
Sample : N1693-19
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
BGZQ2

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/01/2022
Supervised By :Ankita Jodhani 03/04/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 01 09:35:39 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021522CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Feb 16 05:21:54 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

