

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030222\
 Data File : PR053737.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Mar 2022 08:40
 Operator : AJ\MA
 Sample : AIBLK68
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 03 03:30:28 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021522CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Feb 16 05:21:54 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.581	2.808	47968339	27442551	24.488	22.821
2)	SA Decachlor...	9.407	7.967	84141212	64016153	51.649	47.712
Target Compounds							
3)	L1 AR-1016-1	4.746f	0.000	386423	0	6.013	N.D. #
4)	L1 AR-1016-2	4.746f	0.000	386423	0	4.242	N.D. #
9)	L2 AR-1221-2	3.894	0.000	752012	0	44.585	N.D. #
10)	L2 AR-1221-3	3.894f	0.000	752012	0	13.722	N.D. #
11)	L3 AR-1232-1	3.894f	0.000	752012	0	16.422	N.D. #
13)	L3 AR-1232-3	4.746f	0.000	386423	0	9.563	N.D. #
16)	L4 AR-1242-1	4.746f	0.000	386423	0	7.235	N.D. #
17)	L4 AR-1242-2	4.746f	0.000	386423	0	5.168	N.D. #
20)	L4 AR-1242-5	5.768	4.730	6677185	2564792	157.084	71.153 #
21)	L5 AR-1248-1	4.746f	0.000	386423	0	9.659	N.D. #
24)	L5 AR-1248-4	5.712	0.000	15677649	0	212.884	N.D. #
25)	L5 AR-1248-5	5.768	4.730	6677185	2564792	92.501	52.007 #
26)	L6 AR-1254-1	5.712	4.730	15677649	2564792	219.446	34.190 #
27)	L6 AR-1254-2	5.858f	0.000	515492	0	4.576	N.D. #
28)	L6 AR-1254-3	6.357f	5.229f	4734518	1277627	40.326	12.391 #
29)	L6 AR-1254-4	6.623	5.581f	6588135	3385893	72.588	51.436 #
30)	L6 AR-1254-5	7.078	0.000	521025	0	5.374	N.D. #
31)	L7 AR-1260-1	6.553f	5.420	5218129	844116	61.804	11.877 #
32)	L7 AR-1260-2	0.000	5.581f	0	3385893	N.D.	41.524 #
33)	L7 AR-1260-3	7.145	0.000	163796	0	2.189	N.D. #
34)	L7 AR-1260-4	7.421	0.000	419926	0	4.729	N.D. #
35)	L7 AR-1260-5	7.722	0.000	313299	0	1.831	N.D. #
36)	L8 AR-1262-1	7.145	0.000	163796	0	1.433	N.D. #
37)	L8 AR-1262-2	7.722	0.000	313299	0	1.524	N.D. #
38)	L8 AR-1262-3	8.061	0.000	742470	0	7.852	N.D. #
39)	L8 AR-1262-4	8.127	0.000	704358	0	12.641	N.D. #
40)	L8 AR-1262-5	8.796f	0.000	739446	0	9.590	N.D. #
41)	L9 AR-1268-1	8.061	0.000	742470	0	3.043	N.D. #
42)	L9 AR-1268-2	8.127	0.000	704358	0	3.142	N.D. #
44)	L9 AR-1268-4	8.796f	0.000	739446	0	8.892	N.D. #
45)	L9 AR-1268-5	9.115	0.000	234333	0	0.377	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030222\
Data File : PR053737.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Mar 2022 08:40
Operator : AJ\MA
Sample : AIBLK68
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 03 03:30:28 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021522CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Feb 16 05:21:54 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#3 AR-1016-1


R.T.: 4.746 min
Delta R.T.: -0.043 min
Response: 386423
Conc: 6.01 ng/ml

Instrument :
ECD_R
ClientSampleId :




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 3.899 min
Response: 0
Conc: N.D.




#4 AR-1016-2

R.T.: 4.746 min
Delta R.T.: -0.064 min
Response: 386423
Conc: 4.24 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T. : 3.917 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 3.894 min
Delta R.T.: 0.009 min
Response: 752012
Conc: 44.58 ng/ml

Instrument :
ECD_R
ClientSampleId :




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 3.117 min
Response: 0
Conc: N.D.




#10 AR-1221-3

R.T.: 3.894 min
Delta R.T.: -0.069 min
Response: 752012
Conc: 13.72 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 3.191 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 3.894 min
Delta R.T.: -0.069 min
Response: 752012
Conc: 16.42 ng/ml

Instrument :
ECD_R
ClientSampleId :




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 3.190 min
Response: 0
Conc: N.D.




#13 AR-1232-3

R.T.: 4.746 min
Delta R.T.: -0.065 min
Response: 386423
Conc: 9.56 ng/ml



#13 AR-1232-3


R.T.: 0.000 min
Exp R.T. : 4.088 min
Response: 0
Conc: N.D.



#16 AR-1242-1


R.T.: 4.746 min
Delta R.T.: -0.043 min
Response: 386423
Conc: 7.23 ng/ml

Instrument :
ECD_R
ClientSampleId :




#16 AR-1242-1

R.T.: 0.000 min
Exp R.T. : 3.899 min
Response: 0
Conc: N.D.




#17 AR-1242-2

R.T.: 4.746 min
Delta R.T.: -0.064 min
Response: 386423
Conc: 5.17 ng/ml



#17 AR-1242-2


R.T.: 0.000 min
Exp R.T. : 3.916 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 5.768 min
Delta R.T.: 0.010 min
Response: 6677185
Conc: 157.08 ng/ml

Instrument :
ECD_R
ClientSampleId :




#20 AR-1242-5

R.T.: 4.730 min
Delta R.T.: 0.018 min
Response: 2564792
Conc: 71.15 ng/ml




#21 AR-1248-1

R.T.: 4.746 min
Delta R.T.: -0.043 min
Response: 386423
Conc: 9.66 ng/ml



#21 AR-1248-1


R.T.: 0.000 min
Exp R.T.: 3.898 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 5.712 min
Delta R.T.: -0.008 min
Response: 15677649
Conc: 212.88 ng/ml

Instrument :
ECD_R
ClientSampleId :




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 4.351 min
Response: 0
Conc: N.D.




#25 AR-1248-5

R.T.: 5.768 min
Delta R.T.: 0.009 min
Response: 6677185
Conc: 92.50 ng/ml



#25 AR-1248-5


R.T.: 4.730 min
Delta R.T.: -0.023 min
Response: 2564792
Conc: 52.01 ng/ml



#26 AR-1254-1


R.T.: 5.712 min
Delta R.T.: 0.024 min
Response: 15677649
Conc: 219.45 ng/ml

Instrument :
ECD_R
ClientSampleId :




#26 AR-1254-1

R.T.: 4.730 min
Delta R.T.: 0.018 min
Response: 2564792
Conc: 34.19 ng/ml




#27 AR-1254-2

R.T.: 5.858 min
Delta R.T.: -0.064 min
Response: 515492
Conc: 4.58 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T.: 4.869 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 6.357 min
Delta R.T.: 0.042 min
Response: 4734518
Conc: 40.33 ng/ml

Instrument :
ECD_R
ClientSampleId :




#28 AR-1254-3

R.T.: 5.229 min
Delta R.T.: -0.051 min
Response: 1277627
Conc: 12.39 ng/ml




#29 AR-1254-4

R.T.: 6.623 min
Delta R.T.: -0.002 min
Response: 6588135
Conc: 72.59 ng/ml



#29 AR-1254-4


R.T.: 5.581 min
Delta R.T.: 0.057 min
Response: 3385893
Conc: 51.44 ng/ml



#30 AR-1254-5


R.T.: 7.078 min
Delta R.T.: 0.002 min
Response: 521025
Conc: 5.37 ng/ml

Instrument :
ECD_R
ClientSampleId :




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 5.959 min
Response: 0
Conc: N.D.




#31 AR-1260-1

R.T.: 6.553 min
Delta R.T.: 0.063 min
Response: 5218129
Conc: 61.80 ng/ml



#31 AR-1260-1


R.T.: 5.420 min
Delta R.T.: 0.000 min
Response: 844116
Conc: 11.88 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T.: 6.770 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#32 AR-1260-2

R.T.: 5.581 min
Delta R.T.: -0.042 min
Response: 3385893
Conc: 41.52 ng/ml




#33 AR-1260-3

R.T.: 7.145 min
Delta R.T.: -0.010 min
Response: 163796
Conc: 2.19 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T.: 5.776 min
Response: 0
Conc: N.D.



#34 AR-1260-4


R.T.: 7.421 min
Delta R.T.: 0.020 min
Response: 419926
Conc: 4.73 ng/ml

Instrument :
ECD_R
ClientSampleId :




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 6.274 min
Response: 0
Conc: N.D.




#35 AR-1260-5

R.T.: 7.722 min
Delta R.T.: -0.016 min
Response: 313299
Conc: 1.83 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T. : 6.539 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 7.145 min
Delta R.T.: -0.009 min
Response: 163796
Conc: 1.43 ng/ml

Instrument :
ECD_R
ClientSampleId :




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T. : 5.959 min
Response: 0
Conc: N.D.




#37 AR-1262-2

R.T.: 7.722 min
Delta R.T.: -0.014 min
Response: 313299
Conc: 1.52 ng/ml



#37 AR-1262-2


R.T.: 0.000 min
Exp R.T. : 6.537 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 8.061 min
Delta R.T.: 0.029 min
Response: 742470
Conc: 7.85 ng/ml

Instrument :
ECD_R
ClientSampleId :




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 6.837 min
Response: 0
Conc: N.D.




#39 AR-1262-4

R.T.: 8.127 min
Delta R.T.: 0.010 min
Response: 704358
Conc: 12.64 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T. : 6.904 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 8.796 min
Delta R.T.: 0.069 min
Response: 739446
Conc: 9.59 ng/ml

Instrument :
ECD_R
ClientSampleId :




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T. : 7.436 min
Response: 0
Conc: N.D.




#41 AR-1268-1

R.T.: 8.061 min
Delta R.T.: 0.029 min
Response: 742470
Conc: 3.04 ng/ml



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T. : 6.836 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 8.127 min
Delta R.T.: 0.008 min
Response: 704358
Conc: 3.14 ng/ml

Instrument :
ECD_R
ClientSampleId :




#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 6.906 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 8.796 min
Delta R.T.: 0.069 min
Response: 739446
Conc: 8.89 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T. : 7.437 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.115 min
Delta R.T.: 0.022 min
Response: 234333
Conc: 0.38 ng/ml

Instrument :
ECD_R
ClientSampleId :



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 7.727 min
Response: 0
Conc: N.D.