

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030222\
 Data File : PR053752.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Mar 2022 12:34
 Operator : AJ\MA
 Sample : N1693-11
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 BGZP3

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/04/2022
 Supervised By :Ankita Jodhani 03/04/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 03 06:34:03 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021522CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Feb 16 05:21:54 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.564	2.811	31251688	23139690	15.954	19.243
2) SA Decachlor...	9.374	7.955	59169598	42496258	36.321	31.673
Target Compounds						
26) L6 AR-1254-1	5.682	4.708	7272292	8060952	101.793	107.456
27) L6 AR-1254-2	5.919	4.864	13025546	5357231	115.632	81.496 #
28) L6 AR-1254-3	6.313	5.276	6227990	6105645	53.046m	59.217
29) L6 AR-1254-4	6.618	5.518	4700894	3824934	51.794	58.105
30) L6 AR-1254-5	7.069	5.955	8730387	7869898	90.055	83.626

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030222\
Data File : PR053752.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Mar 2022 12:34
Operator : AJ\MA
Sample : N1693-11
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
BGZP3

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/04/2022
Supervised By :Ankita Jodhani 03/04/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 03 06:34:03 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021522CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Feb 16 05:21:54 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

