

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030524\
 Data File : PR065823.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Mar 2024 19:44
 Operator : AJ\MA
 Sample : AR1254ICC100
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12541202

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 06 03:25:56 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR030524CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 06 03:25:46 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.664	2.925	32718457	37331924	5.258	5.215
2) SA Decachlor...	9.541	8.178	17447153	30734603	10.984	10.447
Target Compounds						
26) L6 AR-1254-1	5.787	4.874	20132301	37191128	109.738	106.115
27) L6 AR-1254-2	6.023	5.033	28307206	32860139	109.060	106.543
28) L6 AR-1254-3	6.413	5.456	26036327	49157334	108.140	104.842
29) L6 AR-1254-4	6.725	5.702	13904061	24560966	105.743	104.689
30) L6 AR-1254-5	7.177	6.146	17090908	41323854	105.379	104.158

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030524\
Data File : PR065823.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Mar 2024 19:44
Operator : AJ\MA
Sample : AR1254ICC100
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12541202

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 03:25:56 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR030524CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 06 03:25:46 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

