Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030524\
Data File : PR065833.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Mar 2024 22:11
Operator : AJ\MA
Sample : AR1268ICC100
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12681212

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/06/2024
Supervised By :Ankita Jodhani 03/06/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 04:15:44 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR030524CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 06 04:15:37 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.664	2.924	32955906	36902828	5.234	5.195
2)	SA Decachlor...	9.543	8.178	36285151	69388235	10.407	10.440
	Target Compounds						
41)	L9 AR-1268-1	8.141	7.033	41624054	83498159	105.260m	103.826
42)	L9 AR-1268-2	8.232	7.102	36103286	74357993	104.269	103.698
43)	L9 AR-1268-3	8.447	7.325	33211902	67901800	104.637	103.366
44)	L9 AR-1268-4	8.850	7.637	10520052	24472238	104.857	104.040
45)	L9 AR-1268-5	9.233	7.933	84229939	179.2E6	102.228	102.906

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030524\
Data File : PR065833.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Mar 2024 22:11
Operator : AJ\MA
Sample : AR1268ICC100
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12681212

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/06/2024
Supervised By :Ankita Jodhani 03/06/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 04:15:44 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR030524CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 06 04:15:37 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm