

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030623\
 Data File : PR060067.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Mar 2023 19:28
 Operator : YP\AJ
 Sample : 01805-11MS
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 EYFJ4MS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/07/2023
 Supervised By :Ankita Jodhani 03/07/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 06 20:56:13 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022823CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 28 05:24:37 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	2.904	73560964	98322497	21.836	21.682
2) SA Decachlor...	9.659	8.137	89143469	141.7E6	37.826	39.207
Target Compounds						
3) L1 AR-1016-1	4.923	4.015	43899189	66534352	467.265	484.703
4) L1 AR-1016-2	4.946	4.033	60371393	92547065	447.536	452.886
5) L1 AR-1016-3	5.010	4.210	38023011	47451918	435.849	451.574
6) L1 AR-1016-4	5.113	4.260	31318028	39948068	451.658m	465.208
7) L1 AR-1016-5	5.432	4.476	25667697	52388149	380.371m	468.245
31) L7 AR-1260-1	6.653	5.561	51515232	104.1E6	418.751	455.012
32) L7 AR-1260-2	6.938	5.766	66542601	133.8E6	473.646	497.589
33) L7 AR-1260-3	7.330	5.923	38296836	113.8E6	353.590	449.062 #
34) L7 AR-1260-4	7.575	6.428	46813245	80081643	379.494	380.739
35) L7 AR-1260-5	7.915	6.693	83387171	184.8E6	362.392	388.897

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030623\
Data File : PR060067.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Mar 2023 19:28
Operator : YP\AJ
Sample : 01805-11MS
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
EYFJ4MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/07/2023
Supervised By :Ankita Jodhani 03/07/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 20:56:13 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022823CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 28 05:24:37 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

