

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031122\
 Data File : PR053964.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Mar 2022 21:23
 Operator : AJ\MA
 Sample : N1645-08
 Misc : AR1254 LOQ 50 PPB
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 LOQ-WATER-02-QT1-2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 11 04:07:30 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR031122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 10 01:38:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.558	2.806	43153301	26895862	22.668	21.381
2) SA Decachlor...	9.362	7.951	41562386	30949988	24.791	23.669
Target Compounds						
26) L6 AR-1254-1	5.676	4.703	4076167	4106408	61.793	58.411
27) L6 AR-1254-2	5.910	4.860	6488327	3676194	59.530m	56.649
28) L6 AR-1254-3	6.302	5.271	6789719	5432405	58.712m	55.784
29) L6 AR-1254-4	6.611	5.517	4943155	2855403	54.513m	49.353
30) L6 AR-1254-5	7.062	5.951	5148454	4363959	57.065	47.609

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031122\
Data File : PR053964.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Mar 2022 21:23
Operator : AJ\MA
Sample : N1645-08
Misc : AR1254 LOQ 50 PPB
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
LOQ-WATER-02-QT1-2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 04:07:30 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR031122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 10 01:38:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

