

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031819\
 Data File : PR036280.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Mar 2019 03:34 pm
 Operator : SM\SJ
 Sample : K1847-15
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 SUP-7B-C-2-031119

Manual Integrations
 APPROVED

Sohil
 3/19/2019 9:26:27 AM

Integration File signal 1: sample.E
 Integration File signal 2: events2.e
 Quant Time: Mar 18 15:47:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR031519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Mar 16 00:14:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal #1 Phase : Rxi-1ms Signal #2 Phase:
 Signal #1 Info : 20M x 0.18mm x 0. Signal #2 Info :

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	4.405	3.478	28221093	82627420	15.293	13.001
2) SA Decachlor...	10.066	8.350	20578195	64718567	14.090m	11.341

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031819\
Data File : PR036280.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2019 03:34 pm
Operator : SM\SJ
Sample : K1847-15
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
SUP-7B-C-2-031119


Manual Integrations
APPROVED

Sohil
3/19/2019 9:26:27 AM

Integration File signal 1: sample.E
Integration File signal 2: events2.e
Quant Time: Mar 18 15:47:22 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR031519.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Mar 16 00:14:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal #1 Phase : Rxi-1ms Signal #2 Phase:
Signal #1 Info : 20M x 0.18mm x 0. Signal #2 Info :






#1 Tetrachloro-m-xylene

R.T.: 4.405 min
Delta R.T.: 0.000 min
Response: 28221093
Conc: 15.29 ng/ml

Instrument :
ECD_R
ClientSampleId :
SUP-7B-C-2-031119


Manual Integrations
APPROVED

Sohil
3/19/2019 9:26:27 AM




#1 Tetrachloro-m-xylene

R.T.: 3.478 min
Delta R.T.: 0.000 min
Response: 82627420
Conc: 13.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.066 min
Delta R.T.: 0.000 min
Response: 20578195
Conc: 14.09 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.350 min
Delta R.T.: 0.000 min
Response: 64718567
Conc: 11.34 ng/ml