

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031819\
 Data File : PR036288.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Mar 2019 06:01 pm
 Operator : SM\SJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 I.BLK

Manual Integrations
 APPROVED

Sohil
 3/19/2019 9:26:30 AM

Integration File signal 1: sample.E
 Integration File signal 2: events2.e
 Quant Time: Mar 19 00:35:36 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR031519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Mar 16 00:14:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal #1 Phase : Rxi-1ms Signal #2 Phase:
 Signal #1 Info : 20M x 0.18mm x 0. Signal #2 Info :

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.407	3.477	34451231	100.3E6	18.669	15.777
2)	SA Decachlor...	10.068	8.351	30810590	94524946	21.097m	16.564

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031819\
Data File : PR036288.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2019 06:01 pm
Operator : SM\SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Sohil
3/19/2019 9:26:30 AM

Integration File signal 1: sample.E
Integration File signal 2: events2.e
Quant Time: Mar 19 00:35:36 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR031519.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Mar 16 00:14:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal #1 Phase : Rxi-1ms Signal #2 Phase:
Signal #1 Info : 20M x 0.18mm x 0. Signal #2 Info :






#1 Tetrachloro-m-xylene

R.T.: 4.407 min
Delta R.T.: 0.002 min
Response: 34451231
Conc: 18.67 ng/ml

Instrument :
ECD_R
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Sohil
3/19/2019 9:26:30 AM




#1 Tetrachloro-m-xylene

R.T.: 3.477 min
Delta R.T.: -0.001 min
Response: 100269853
Conc: 15.78 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.068 min
Delta R.T.: 0.002 min
Response: 30810590
Conc: 21.10 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.351 min
Delta R.T.: 0.000 min
Response: 94524946
Conc: 16.56 ng/ml