

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031819\
 Data File : PR036295.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Mar 2019 08:03 pm
 Operator : SM\SJ
 Sample : K1914-01RE
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 CONCRETE-PILERE

Manual Integrations
 APPROVED

Sohil
 3/19/2019 9:26:32 AM

Integration File signal 1: sample.E
 Integration File signal 2: events2.e
 Quant Time: Mar 19 00:39:42 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR031519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Mar 16 00:14:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal #1 Phase : Rxi-1ms Signal #2 Phase:
 Signal #1 Info : 20M x 0.18mm x 0. Signal #2 Info :

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.406	3.479	18195852	51180979	9.860m	8.053
2) SA Decachlor...	10.062	8.347	16148174	50574861	11.057	8.862

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031819\
Data File : PR036295.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2019 08:03 pm
Operator : SM\SJ
Sample : K1914-01RE
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
CONCRETE-PILERE


Manual Integrations
APPROVED

Sohil
3/19/2019 9:26:32 AM

Integration File signal 1: sample.E
Integration File signal 2: events2.e
Quant Time: Mar 19 00:39:42 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR031519.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Mar 16 00:14:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal #1 Phase : Rxi-1ms Signal #2 Phase:
Signal #1 Info : 20M x 0.18mm x 0. Signal #2 Info :






#1 Tetrachloro-m-xylene

R.T.: 4.406 min
Delta R.T.: 0.000 min
Response: 18195852
Conc: 9.86 ng/ml m

Instrument :
ECD_R
ClientSampleId :
CONCRETE-PILERE


Manual Integrations
APPROVED

Sohil
3/19/2019 9:26:32 AM




#1 Tetrachloro-m-xylene

R.T.: 3.479 min
Delta R.T.: 0.000 min
Response: 51180979
Conc: 8.05 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.062 min
Delta R.T.: -0.005 min
Response: 16148174
Conc: 11.06 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.347 min
Delta R.T.: -0.003 min
Response: 50574861
Conc: 8.86 ng/ml