

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031923\
 Data File : PR060346.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Mar 2023 23:45
 Operator : AJ\MA
 Sample : 01960-17
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 EYFG2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/20/2023
 Supervised By :Ankita Jodhani 03/20/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 20 03:53:57 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022823CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 28 05:24:37 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	2.903	69704236	99361842	20.691	21.912
2) SA Decachlor...	9.657	8.131	134.8E6	210.7E6	57.214	58.301
Target Compounds						
16) L4 AR-1242-1	4.922	4.013	1846643	2647316	23.501m	23.195
17) L4 AR-1242-2	4.944	4.031	3079873	3348687	27.467m	20.015 #
18) L4 AR-1242-3	5.011	4.208	2061772	1691482	28.382m	19.554 #
19) L4 AR-1242-4	5.111	4.300	1690564	2496393	29.372m	29.126
20) L4 AR-1242-5	5.907	4.842	1990702	4108887	36.839m	39.790

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031923\
Data File : PR060346.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Mar 2023 23:45
Operator : AJ\MA
Sample : 01960-17
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
EYFG2

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/20/2023
Supervised By :Ankita Jodhani 03/20/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 20 03:53:57 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022823CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 28 05:24:37 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

