

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR032019\
 Data File : PR036413.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Mar 2019 17:55
 Operator : SM\SJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 21 07:34:53 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR032019.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 20 15:19:33 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.401	3.472	35901147	99406039	23.720	24.243
2) SA Decachlor...	10.057	8.340	35083502	106.8E6	20.016	19.306
Target Compounds						
32) L7 AR-1260-2	0.000	6.151	0	3288745	N.D.	7.389 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR032019\
Data File : PR036413.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Mar 2019 17:55
Operator : SM\SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 21 07:34:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR032019.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 20 15:19:33 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.401 min
Delta R.T.: 0.000 min
Response: 35901147
Conc: 23.72 ng/ml

Instrument :
ECD_R
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.472 min
Delta R.T.: -0.001 min
Response: 99406039
Conc: 24.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.057 min
Delta R.T.: 0.000 min
Response: 35083502
Conc: 20.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.340 min
Delta R.T.: 0.000 min
Response: 106806046
Conc: 19.31 ng/ml



#32 AR-1260-2

R.T.: 0.000 min
Exp R.T.: 7.409 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :



#32 AR-1260-2

R.T.: 6.151 min
Delta R.T.: -0.008 min
Response: 3288745
Conc: 7.39 ng/ml