

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR032023\
 Data File : PR060409.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Mar 2023 17:35
 Operator : AJ\MA
 Sample : 01960-08
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 EYFG3

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/21/2023
 Supervised By :Ankita Jodhani 03/21/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 21 02:10:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022823CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 28 05:24:37 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.689	2.903	60787443	85960672	18.044	18.956
2) SA Decachlor...	9.656	8.131	79135241	118.8E6	33.579	32.862
Target Compounds						
16) L4 AR-1242-1	4.922	4.012	1003132	1530158	12.766m	13.407
17) L4 AR-1242-2	4.946	4.030	2251873	2031522	20.082m	12.142 #
18) L4 AR-1242-3	5.011	4.208	2306812	1210820	31.756m	13.997 #
19) L4 AR-1242-4	5.111	4.300	1829324	2115143	31.783m	24.678
20) L4 AR-1242-5	5.906	4.843	2119560	4111551	39.223m	39.815

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR032023\
Data File : PR060409.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Mar 2023 17:35
Operator : AJ\MA
Sample : 01960-08
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
EYFG3

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/21/2023
Supervised By :Ankita Jodhani 03/21/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 21 02:10:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022823CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 28 05:24:37 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

