Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR032123\
Data File : PR060507.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Mar 2023 17:36
Operator : AJ\MA
Sample : 01991-01
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
E0277

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/22/2023
Supervised By :Ankita Jodhani 03/22/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 22 09:53:59 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022823CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 28 05:24:37 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.688	2.902	59964046	84186056	17.800	18.565
2)	SA Decachlor...	9.657	8.132	111.2E6	169.5E6	47.184	46.909
	Target Compounds						
21)	L5 AR-1248-1	4.922	4.012	5033243	6767555	83.059m	77.753
22)	L5 AR-1248-2	5.217	4.259	6427349	7799681	79.159m	60.421
23)	L5 AR-1248-3	5.432	4.300	6611114	7964013	70.382	61.259
24)	L5 AR-1248-4	5.867	4.474	6971448	9061969	72.536m	57.428
25)	L5 AR-1248-5	5.907	4.883	7527515	9707430	81.124m	68.126

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR032123\
Data File : PR060507.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Mar 2023 17:36
Operator : AJ\MA
Sample : 01991-01
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
E0277

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/22/2023
Supervised By :Ankita Jodhani 03/22/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 22 09:53:59 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022823CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 28 05:24:37 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm