

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR032123\
 Data File : PR060560.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Mar 2023 06:36
 Operator : AJ\MA
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12543317

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/22/2023
 Supervised By :Ankita Jodhani 03/22/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 22 06:56:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022823CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 28 05:24:37 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	2.902	57637251	77996385	17.109	17.200
2) SA Decachlor...	9.656	8.130	101.1E6	150.5E6	42.903	41.629
Target Compounds						
26) L6 AR-1254-1	5.838	4.840	42801369	93727279	427.007	403.031
27) L6 AR-1254-2	6.077	4.999	63407471	74000566	417.609	368.571
28) L6 AR-1254-3	6.472	5.418	67908183	125.0E6	437.013	390.212m
29) L6 AR-1254-4	6.786	5.664	40964946	84372102	371.813	470.378 #
30) L6 AR-1254-5	7.243	6.107	54934014	117.4E6	447.779	424.426

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR032123\
Data File : PR060560.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Mar 2023 06:36
Operator : AJ\MA
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12543317

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/22/2023
Supervised By :Ankita Jodhani 03/22/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 22 06:56:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022823CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 28 05:24:37 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

