

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR032223\
 Data File : PR060573.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Mar 2023 12:19
 Operator : AJ\MA
 Sample : 01991-12MSD
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 E0286MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/23/2023
 Supervised By :Ankita Jodhani 03/23/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 22 23:49:38 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022823CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 28 05:24:37 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.689	2.902	74248227	98967298	22.040	21.825
2) SA Decachlor...	9.657	8.129	108.1E6	168.8E6	45.886	46.700
Target Compounds						
3) L1 AR-1016-1	4.922	4.011	44678944	64828908	475.564	472.279
4) L1 AR-1016-2	4.945	4.029	63335174	90080348	469.506	440.814
5) L1 AR-1016-3	5.010	4.206	44724194	46780175	512.663	445.181
6) L1 AR-1016-4	5.113	4.257	34335926	39562171	495.182	460.714
7) L1 AR-1016-5	5.431	4.473	29955274	44349480	443.909	396.395
31) L7 AR-1260-1	6.653	5.558	53613844	93942145	435.809	410.773
32) L7 AR-1260-2	6.927	5.763	11405.1E6	134.1E6	81180.405	498.429 #
33) L7 AR-1260-3	7.328	5.919	48964975	121.8E6	452.088	480.690
34) L7 AR-1260-4	7.573	6.423	45806755	95681594	371.335m	454.907
35) L7 AR-1260-5	7.913	6.688	98813706	244.0E6	429.435	513.361

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR032223\
Data File : PR060573.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Mar 2023 12:19
Operator : AJ\MA
Sample : 01991-12MSD
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
E0286MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/23/2023
Supervised By :Ankita Jodhani 03/23/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 22 23:49:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022823CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 28 05:24:37 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

