

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR032619\
 Data File : PR036692.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Mar 2019 10:17
 Operator : SM\SJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 27 04:30:25 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR032319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 25 03:18:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.404	3.472	37543337	106.6E6	22.289	20.434
2)	SA Decachlor...	10.050	8.335	33611636	102.5E6	22.049	19.722
Target Compounds							
31)	L7 AR-1260-1	7.169	0.000	23078128	0	261.300	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR032619\
Data File : PR036692.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Mar 2019 10:17
Operator : SM\SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 04:30:25 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR032319.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 25 03:18:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.404 min
Delta R.T.: 0.008 min
Response: 37543337
Conc: 22.29 ng/ml

Instrument :
ECD_R
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.472 min
Delta R.T.: 0.004 min
Response: 106573727
Conc: 20.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.050 min
Delta R.T.: 0.004 min
Response: 33611636
Conc: 22.05 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.335 min
Delta R.T.: 0.002 min
Response: 102478708
Conc: 19.72 ng/ml



#31 AR-1260-1

R.T.: 7.169 min
Delta R.T.: 0.022 min
Response: 23078128
Conc: 261.30 ng/ml

Instrument :
ECD_R
ClientSampleId :
I.BLK



#31 AR-1260-1

R.T.: 0.000 min
Exp R.T. : 5.968 min
Response: 0
Conc: N.D.