

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR040221\
 Data File : PR049753.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Apr 2021 02:34
 Operator : DD\AJ
 Sample : M1458-07
 Misc : AR1232 LOD 40 PPB
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 LOD-MDL-WATER-01-QT1-2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 05 01:54:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Apr 03 08:12:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.636	3.820	57169495	88638173	14.642	13.366
2) SA Decachlor...	10.512	8.862	70210265	85937172	15.802	11.288 #
Target Compounds						
11) L3 AR-1232-1	5.015	4.197	5598678	7635701	72.455	60.566
12) L3 AR-1232-2	5.549	4.929	3190413	7304318	75.277	54.990 #
13) L3 AR-1232-3	5.841	5.106	6494467	3783368	76.230	54.415 #
14) L3 AR-1232-4	6.004	5.189	3480804	3292269	81.672	56.003 #
15) L3 AR-1232-5	6.090	5.361	2780178	3829328	84.069	58.005 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR040221\
Data File : PR049753.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Apr 2021 02:34
Operator : DD\AJ
Sample : M1458-07
Misc : AR1232 LOD 40 PPB
ALS Vial : 54 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 05 01:54:50 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040221.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Apr 03 08:12:17 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

