

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR040323\
 Data File : PR060643.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Apr 2023 21:17
 Operator : YP\AJ
 Sample : AR1262ICC200
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 04 01:31:06 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040323CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 01:29:26 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.688	2.901	34365424	46577892	10.152	9.786
2)	SA Decachlor...	9.660	8.129	67675486	96927871	21.795	20.504
Target Compounds							
36)	L8 AR-1262-1	7.331	6.149	42853797	71498459	214.968	204.090
37)	L8 AR-1262-2	7.915	6.421	74532844	64794141	211.020	203.911
38)	L8 AR-1262-3	8.227	6.991	52885204	51628350	214.695	206.271
39)	L8 AR-1262-4	8.314	7.058	40434256	98804950	213.889	203.039
40)	L8 AR-1262-5	8.948	7.595	29361695	45515813	213.344	204.768



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR040323\
Data File : PR060643.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Apr 2023 21:17
Operator : YP\AJ
Sample : AR1262ICC200
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 04 01:31:06 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040323CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 01:29:26 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#36 AR-1262-1


R.T.: 7.331 min
Delta R.T.: 0.000 min
Response: 42853797
Conc: 214.97 ng/ml

Instrument :
ECD_R
ClientSampleId :




#36 AR-1262-1

R.T.: 6.149 min
Delta R.T.: 0.000 min
Response: 71498459
Conc: 204.09 ng/ml




#37 AR-1262-2

R.T.: 7.915 min
Delta R.T.: 0.000 min
Response: 74532844
Conc: 211.02 ng/ml



#37 AR-1262-2


R.T.: 6.421 min
Delta R.T.: 0.000 min
Response: 64794141
Conc: 203.91 ng/ml



#38 AR-1262-3


R.T.: 8.227 min
Delta R.T.: 0.000 min
Response: 52885204
Conc: 214.70 ng/ml

Instrument :
ECD_R
ClientSampleId :




#38 AR-1262-3

R.T.: 6.991 min
Delta R.T.: 0.000 min
Response: 51628350
Conc: 206.27 ng/ml




#39 AR-1262-4

R.T.: 8.314 min
Delta R.T.: 0.002 min
Response: 40434256
Conc: 213.89 ng/ml



#39 AR-1262-4


R.T.: 7.058 min
Delta R.T.: 0.000 min
Response: 98804950
Conc: 203.04 ng/ml



#40 AR-1262-5

R.T.: 8.948 min
Delta R.T.: 0.000 min
Response: 29361695
Conc: 213.34 ng/ml

Instrument :
ECD_R
ClientSampleId :



#40 AR-1262-5

R.T.: 7.595 min
Delta R.T.: 0.000 min
Response: 45515813
Conc: 204.77 ng/ml