

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR040521\
 Data File : PR049777.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Apr 2021 12:53
 Operator : DD\AJ
 Sample : PR040521ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 06 01:45:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 06 01:34:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	4.289	4.604 E6	-7.3	104	0.00
2 SA Decachlorobiphenyl	6.443	6.547 E6	-1.6	101	0.00
3 L1 AR-1016-1	213.008	216.530 E3	-1.7	100	0.00
4 L1 AR-1016-2	303.909	304.935 E3	-0.3	98	0.00
5 L1 AR-1016-3	182.502	182.454 E3	0.0	98	0.00
6 L1 AR-1016-4	152.223	154.393 E3	-1.4	100	0.00
7 L1 AR-1016-5	155.644	155.914 E3	-0.2	100	0.00
31 L7 AR-1260-1	283.093	283.654 E3	-0.2	100	0.00
32 L7 AR-1260-2	344.608	348.463 E3	-1.1	101	0.00
33 L7 AR-1260-3	265.522	269.487 E3	-1.5	101	0.00
34 L7 AR-1260-4	312.385	316.961 E3	-1.5	100	0.00
35 L7 AR-1260-5	651.178	672.427 E3	-3.3	101	0.00

Signal #2

1 SA Tetrachloro-m-xylene	7.191	7.650 E6	-6.4	103	0.00
2 SA Decachlorobiphenyl	8.159	8.474 E6	-3.9	102	0.00
3 L1 AR-1016-1	247.633	254.827 E3	-2.9	102	0.00
4 L1 AR-1016-2	371.566	378.802 E3	-1.9	100	0.00
5 L1 AR-1016-3	197.306	198.614 E3	-0.7	100	0.00
6 L1 AR-1016-4	158.117	154.906 E3	2.0	99	0.00
7 L1 AR-1016-5	205.633	204.383 E3	0.6	100	0.00
31 L7 AR-1260-1	396.806	397.888 E3	-0.3	100	0.00
32 L7 AR-1260-2	480.983	487.614 E3	-1.4	101	0.00
33 L7 AR-1260-3	459.303	466.392 E3	-1.5	100	0.00
34 L7 AR-1260-4	390.015	400.017 E3	-2.6	101	0.00
35 L7 AR-1260-5	953.212	1005.502 E3	-5.5	103	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	47.999	0.000 E3	100.0#	0#	-4.85#
9 L2 AR-1221-2	37.012	0.000 E3	100.0#	0#	-4.94#
10 L2 AR-1221-3	114.648	0.000 E3	100.0#	0#	-5.02#
11 L3 AR-1232-1	93.459	0.000 E3	100.0#	0#	-5.02#
12 L3 AR-1232-2	57.278	0.000 E3	100.0#	0#	-5.55#
13 L3 AR-1232-3	117.225	0.000 E3	100.0#	0#	-5.84#
14 L3 AR-1232-4	59.740	0.000 E3	100.0#	0#	-6.00#
15 L3 AR-1232-5	47.333	0.000 E3	100.0#	0#	-6.09#
16 L4 AR-1242-1	159.555	0.000 E3	100.0#	0#	-5.82#
17 L4 AR-1242-2	225.464	0.000 E3	100.0#	0#	-5.84#
18 L4 AR-1242-3	136.133	0.000 E3	100.0#	0#	-5.90#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR040521\
 Data File : PR049777.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Apr 2021 12:53
 Operator : DD\AJ
 Sample : PR040521ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 06 01:45:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 06 01:34:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
19 L4 AR-1242-4	114.457	0.000 E3	100.0#	0#	-6.00#
20 L4 AR-1242-5	127.927	0.000 E3	100.0#	0#	-6.74#
21 L5 AR-1248-1	115.715	0.000 E3	100.0#	0#	-5.82#
22 L5 AR-1248-2	168.372	0.000 E3	100.0#	0#	-6.09#
23 L5 AR-1248-3	190.067	0.000 E3	100.0#	0#	-6.30#
24 L5 AR-1248-4	225.577	0.000 E3	100.0#	0#	-6.70#
25 L5 AR-1248-5	217.955	0.000 E3	100.0#	0#	-6.74#
26 L6 AR-1254-1	224.436	0.000 E3	100.0#	0#	-6.67#
27 L6 AR-1254-2	350.170	0.000 E3	100.0#	0#	-6.89#
28 L6 AR-1254-3	381.052	0.000 E3	100.0#	0#	-7.26#
29 L6 AR-1254-4	288.390	0.000 E3	100.0#	0#	-7.55#
30 L6 AR-1254-5	311.115	0.000 E3	100.0#	0#	-7.97#
36 L8 AR-1262-1	368.793	0.000 E3	100.0#	0#	-8.04#
37 L8 AR-1262-2	713.712	0.000 E3	100.0#	0#	-8.61#
38 L8 AR-1262-3	467.654	0.000 E3	100.0#	0#	-8.93#
39 L8 AR-1262-4	350.938	0.000 E3	100.0#	0#	-9.03#
40 L8 AR-1262-5	258.912	0.000 E3	100.0#	0#	-9.73#
41 L9 AR-1268-1	857.003	0.000 E3	100.0#	0#	-8.93#
42 L9 AR-1268-2	787.507	0.000 E3	100.0#	0#	-9.03#
43 L9 AR-1268-3	675.186	0.000 E3	100.0#	0#	-9.27#
44 L9 AR-1268-4	292.942	0.000 E3	100.0#	0#	-9.73#
45 L9 AR-1268-5	2.421	0.000 E6	100.0#	0#	-10.16#

Signal #2

8 L2 AR-1221-1	69.814	0.000 E3	100.0#	0#	-4.04#
9 L2 AR-1221-2	53.779	0.000 E3	100.0#	0#	-4.12#
10 L2 AR-1221-3	170.205	0.000 E3	100.0#	0#	-4.20#
11 L3 AR-1232-1	138.075	0.000 E3	100.0#	0#	-4.20#
12 L3 AR-1232-2	145.427	0.000 E3	100.0#	0#	-4.93#
13 L3 AR-1232-3	76.915	0.000 E3	100.0#	0#	-5.11#
14 L3 AR-1232-4	67.597	0.000 E3	100.0#	0#	-5.19#
15 L3 AR-1232-5	74.701	0.000 E3	100.0#	0#	-5.36#
16 L4 AR-1242-1	187.724	0.000 E3	100.0#	0#	-4.91#
17 L4 AR-1242-2	279.380	0.000 E3	100.0#	0#	-4.93#
18 L4 AR-1242-3	148.107	0.000 E3	100.0#	0#	-5.11#
19 L4 AR-1242-4	139.807	0.000 E3	100.0#	0#	-5.19#
20 L4 AR-1242-5	182.443	0.000 E3	100.0#	0#	-5.71#
21 L5 AR-1248-1	138.804	0.000 E3	100.0#	0#	-4.91#
22 L5 AR-1248-2	194.536	0.000 E3	100.0#	0#	-5.15#
23 L5 AR-1248-3	205.800	0.000 E3	100.0#	0#	-5.19#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR040521\
 Data File : PR049777.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Apr 2021 12:53
 Operator : DD\AJ
 Sample : PR040521ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 06 01:45:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 06 01:34:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	252.298	0.000 E3	100.0#	0#	-5.36#
25 L5 AR-1248-5	263.077	0.000 E3	100.0#	0#	-5.75#
26 L6 AR-1254-1	399.426	0.000 E3	100.0#	0#	-5.71#
27 L6 AR-1254-2	343.019	0.000 E3	100.0#	0#	-5.86#
28 L6 AR-1254-3	596.333	0.000 E3	100.0#	0#	-6.26#
29 L6 AR-1254-4	351.594	0.000 E3	100.0#	0#	-6.49#
30 L6 AR-1254-5	522.064	0.000 E3	100.0#	0#	-6.91#
36 L8 AR-1262-1	268.870	0.000 E3	100.0#	0#	-6.91#
37 L8 AR-1262-2	1.016	0.000 E6	100.0#	0#	-7.44#
38 L8 AR-1262-3	386.077	0.000 E3	100.0#	0#	-7.73#
39 L8 AR-1262-4	759.690	0.000 E3	100.0#	0#	-7.79#
40 L8 AR-1262-5	331.016	0.000 E3	100.0#	0#	-8.29#
41 L9 AR-1268-1	1.204	0.000 E6	100.0#	0#	-7.73#
42 L9 AR-1268-2	1.155	0.000 E6	100.0#	0#	-7.79#
43 L9 AR-1268-3	954.150	0.000 E3	100.0#	0#	-8.00#
44 L9 AR-1268-4	383.783	0.000 E3	100.0#	0#	-8.29#
45 L9 AR-1268-5	3.199	0.000 E6	100.0#	0#	-8.59#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0