

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR040721\
 Data File : PR049874.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Apr 2021 10:29
 Operator : DD\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12543005

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 08 01:46:52 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040621CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 07 06:22:57 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.636	3.820	77314043	130.8E6	17.379	17.724
2) SA Decachlor...	10.508	8.857	221.7E6	275.2E6	33.589	32.934
Target Compounds						
26) L6 AR-1254-1	6.673	5.709	117.0E6	226.2E6	334.567	345.644
27) L6 AR-1254-2	6.890	5.856	188.1E6	193.2E6	338.364	343.489
28) L6 AR-1254-3	7.260	6.259	212.4E6	339.0E6	338.961	344.405
29) L6 AR-1254-4	7.547	6.486	163.3E6	206.9E6	339.807	342.423
30) L6 AR-1254-5	7.966	6.903	181.9E6	297.9E6	340.263	341.900

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR040721\
Data File : PR049874.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2021 10:29
Operator : DD\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12543005

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 01:46:52 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040621CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 07 06:22:57 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

