

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR040823\
 Data File : PR060764.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Apr 2023 04:32
 Operator : AJ\MA
 Sample : 02213-02
 Misc : AR1248 LOQ 50 PPB
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 LOQ-SOIL-02-QT2-2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 01:11:44 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 07 15:47:54 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.687	2.900	65351976	84811194	21.406	19.928
2) SA Decachlor...	9.657	8.125	65202213	90753971	22.985	20.876
Target Compounds						
21) L5 AR-1248-1	4.922	4.009	4017069	4354849	60.449	50.646
22) L5 AR-1248-2	5.217	4.255	5656279	6337400	63.703	52.706
23) L5 AR-1248-3	5.431	4.296	6355972	6479591	60.371	53.155
24) L5 AR-1248-4	5.867	4.470	6995297	7801979	58.183	51.688
25) L5 AR-1248-5	5.906	4.879	6995195	7979314	59.091	51.914

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR040823\
Data File : PR060764.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2023 04:32
Operator : AJ\MA
Sample : 02213-02
Misc : AR1248 LOQ 50 PPB
ALS Vial : 55 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
LOQ-SOIL-02-QT2-2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 01:11:44 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040823.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 07 15:47:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

