

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042324\
 Data File : PR066363.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Apr 2024 19:17
 Operator : AJ\MA
 Sample : AR1242ICC1600
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12425215

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 22:49:53 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR042324CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 23 22:49:47 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.666	2.897	395.4E6	343.0E6	68.865	69.715
2) SA Decachlor...	9.578	8.132	177.1E6	341.4E6	128.829	131.263
Target Compounds						
16) L4 AR-1242-1	4.885	4.009	160.6E6	170.3E6	1305.219	1317.061
17) L4 AR-1242-2	4.907	4.026	248.7E6	243.3E6	1322.331	1330.552
18) L4 AR-1242-3	4.972	4.204	158.5E6	134.4E6	1267.177	1312.369
19) L4 AR-1242-4	5.074	4.296	128.5E6	128.3E6	1288.083	1266.916
20) L4 AR-1242-5	5.859	4.840	125.7E6	161.6E6	1317.994	1308.504

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042324\
Data File : PR066363.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Apr 2024 19:17
Operator : AJ\MA
Sample : AR1242ICC1600
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12425215

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 22:49:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR042324CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 23 22:49:47 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

